



**Sarada Krishna
Homoeopathic Medical College
Kulasekharam, Kanniyakumari Dist., Tamil Nadu**

DETAILS OF MAINTENANCE

Sl.No	Details
1.	Maintenance policy statement
2.	AMC document for Generator
3.	AMC document for Lift
4.	Structure stability certificate for institution
5.	Fire service license
6.	Building license
7.	Calibration certificates for equipments




PRINCIPAL
SARADA KRISHNA HOMOEOPATHIC MEDICAL COLLEGE
KULASEKHARAM, KANNIYAKUMARI DISTRICT,
TAMIL NADU - 629 151



**Sarada Krishna
Homoeopathic Medical College
Kulasekharam, Kanniyakumari Dist., Tamil Nadu-629 161.**

MAINTENANCE COMMITTEE

Purpose

Purpose of maintenance policy is to carry out the maintenance work in the campus for the enhancement of physical, academic and support systems for teaching and learning. The maintenance committee is responsible for managing the maintenance facilities as the most cost effective manner.

Scope

The scope of the maintenance policy is limited to undertake maintenance of the infrastructure of the campus. Maintenance policy of SKHMC ensures no major breakdowns in the academic and other supporting system.

Policy

The SKHMC maintenance system will include

- Priorities for work requests
- Comprehensive working procedure
- Performance Goals
- Long range plan

The priority system ensures the follows:

- a. Emergency
- b. Urgent
- c. Scheduled Operations and Services

Responsibilities

The maintenance committee has the following members

1. The Chairman
2. Principal as Co-ordinator
3. Maintenance Manager
4. Office Manager
5. Office Superintendent
6. Head Accountant
7. Quality Coordinator
8. Vehicle Supervisor

Procedure


1. Maintenance request by individual units / website.
2. Maintenance Committee scrutinize the maintenance request and put up for approval of the competent authority.
3. After completion of repairs/maintenance cost is settled as per the approved procedure.

The maintenance manager will ensure that the staff implements the maintenance policy as per the procedures. He is responsible for developing the work culture to identify the means of delivery. The SKHMC will contract maintenance services when it is the best interest of the organization to do so. When the employees of the organization have the time and skills to perform the work at hand they will be the first choice to perform the given task.

Maintenance reference to the inspection, modification and repair of buildings, equipments and machines within the organization to ensure they are in good conditions. The good maintenance of facilities is also an economic means of maximizing the full value of these assets.


Co-ordinator




Principal
PRINCIPAL
Sarada Krishna Homoeopathic
Medical College, Kula. . .

14

M3

3837

GOVERNMENT OF TAMILNADU
ELECTRICAL INSPECTORATE

Phone & Fax : 0462 2530343

e-mail : eitn@tn.gov.in
Visit : www.tnei.tn.gov.in

From
The Electrical Inspector,
7-B, 'C' Colony,
Perumalpuram,
Tirunelveli - 627 007.

To
M/s. Sarada Krishna Homoeopathic
Medical College,
Kulasekharam,
Kanyakumari District - 629 161.

Pell 53
12

Letter No. 1963/EI/TIN/Reg 32/2017 Dated :08.06.2017

Sirs,

Sub : Electricity - Generator installations - Voltages upto and inclusive of 415V at the premises of **M/s. Sarada Krishna Homoeopathic Medical College, Kulasekharam, Kanyakumari District - 629 161**- Inspection under Regulation 32 of Central Electricity Authority (Measures relating to Safety and Electric Supply) Regulations, 2010 - Inspected on **27.05.2017** - Permission to Commission the Equipments - Reg.

- Ref : 1) Your Letter No.Nil Dt.: 18.05.2017.
 2) This Office Lr. No.TIN 1963/AEI/TIN/DG/R32/2017-18 Dt:29.05.2017.
 3) Your Letter No.Nil. Dt: 01.06.2017 Recd. on 05.06.2017.

Under Regulation 32 of Central Electricity Authority (Measures relating to Safety and Electric Supply) Regulations, 2010, permission is hereby accorded to commission the Diesel Generator Set bearing:

	<u>Alternator -1</u>	<u>Engine - 1</u>	<u>Alternator -2</u>	<u>Engine - 2</u>
Make	Kirloskar	Kirloskar	Kirloskar	Kirloskar
SLNo.	ES3H009C71098	6H.3532 / 0900092	ES3D016F205012	4H.8901 / 1620894
Rating	125 KVA/415V	156 HP	82.5 KVA/415V	102 HP

CONDITIONS

Subject to the following Conditions:

- 1) The Generator supply should not under any circumstances be run in parallel with the TANGEDCO Grid. Adequate measures should be taken by always maintaining standard change over switches.
- 2) The permission granted now ceases to be valid when the Generator set is shifted from the above service connection.
- 3) The date of commissioning of the Generator set should be intimated to this office, The Superintending Engineer, Kanyakumari / TANGEDCO, The Executive Engineer, Thuckalay TANGEDCO and Chief Electrical Inspector to Government, Chennai-32.
- 4) Tax on Consumption or Sale of Electricity using generator supply should be remitted monthly and returns sent to this office as per Tamil Nadu Tax on Consumption or Sale of Electricity Act, 2003.

Electrical Inspector
Tirunelveli

Copy to: Thiru.Global Technics Solution, 266, Cheranmahadevi Road, Pettai, Tirunelveli.
Copy to: The Superintending Engineer/ TANGEDCO LTD.,
Kanyakumari Electricity Distribution Circle, Kanyakumari.
Copy submitted to: The Chief Electrical Inspector to Govt., Chennai.

Stamp: SARADA KRISHNA HOMOEOPATHIC MEDICAL COLLEGE
KULASEKHARAM
KANYAKUMARI DISTRICT
08 JUN 2017

KIRLOSKAR OIL ENGINES LIMITED
A Kirloskar Group Company



3482 (4)
KOEL
CARE

Date: 18-May-2018

To,

Sarada Krishna Homeopathic Medical College

Kulasekharam Kanyakumari

Kanyakumari - 629161

Kind Attn: Mr. Sanjeev Kumar

Dear Sir/Madam,

It gives us immense pleasure to welcome you to **KOEL BANDHAN** family.

We thank you for choosing **KOEL BANDHAN** and giving us the opportunity to take care of your **KOEL Green** asset. It is our privilege to serve you. We shall ensure that you always avail the best of our services through our 5 promises:

- Proactiveness
- Speed - Response & Restoration
- First Time Right
- One Stop Solution
- Complete Service Ownership

Kindly take a note of the following important details pertaining to your online purchase:

- Engine Serial Number : 4H.8901/1620894
- Instance ID of your **KOEL Green** Genset : 100156313
- Invoice No and Date : 331929001300156 Dt. 18/May/18
- Start date of contract : 17/May/18
- End date of contract : 16/May/19

For any assistance, you can always contact our 24x7 helpdesk @ **8806334433** or Email us on koel.helpdesk@kirloskar.com

Note: Please mention your Genset Instance ID/ Engine Serial Number as mentioned above while contacting our 24x7 Helpdesk to track your genset easily and serve you better.

We once again thank you very much for choosing **KOEL BANDHAN** and assure you of our best services always.

Regards,

Abhaya Naik
(Head - Customer Support)

Enclosure : Scope, Inclusion and Exclusion of **KOEL Bandhan** Service Package.

4



SARADA KRISHNA HOMOEOPATHIC
 MEDICAL COLLEGE HOSPITAL
 12 MAR 2019
 KULASEKHARAM

SARADA KRISHNA HOMOEOPATHIC
 MEDICAL COLLEGE
 KULASEKHARAM
 KANYAKUMARI
 629161
 India

KONE Elevator India Pvt Ltd

NO.1,PANDIAN TOWER,
 MELAPONNAGARAM 5TH
 STREET,
 MADURAI
 625016
 Tel:0452 2361617
 Contact person:S K Kumaresh
 Contact person mob:+91 80560
 41261

Contract No: 41457498

Quotation No. T-0002825756 v1

27.02.2019

KONE Care™ contract

Dear

We wish to express our sincere gratitude to you for continuing to be KONE customer.

Your continued patronage has helped us to provide uninterrupted services to KONE Elevators / Escalators in your premises. It has been our endeavor to ensure a safe and comfortable ride for all users of KONE Equipment in your building.

It is quite some time, since we signed the full set of contract documents; hence it is time for us to renew our agreement towards the scope, terms & conditions of the referred Maintenance Contract. Further, the costs of labour and material have seen a substantial upward movement during the last one year. We are forced to pass on a certain portion of the increased cost to continue maintaining our service routines. Accordingly, we propose a nominal increase over the present contract value.

Enclosed please find the contract document. We would request you to please sign the copies on each page and return to us to enable us, process your instruction as speedily and efficiently as possible.

However should you have any questions regarding our proposal, or require any further information, please do not hesitate to contact us.

Yours sincerely
 KONE Elevator India Pvt Ltd
 S K Kumaresh
 Engineer - Service
 sk.kumaresh@kone.com

SKHMC	
Principal	
Chairman	

For KONE Elevator India Pvt Ltd
 For KONE Elevator India Pvt Ltd

KONE ELEVATOR INDIA PVT. LTD.

Signed For Customer



KONE Care Standard™

Sold to: SARADA KRISHNA
HOMOEOPATHIC
KULASEKHARAM
KANYAKUMARI
TAMIL NADU
629161
India

Invoice to: SARADA KRISHNA
HOMOEOPATHIC MEDICAL
COLLEGE
KULASEKHARAM
629161
KANYAKUMARI

and contractor:

KONE Elevator India Pvt Ltd
NO.1,PANDIAN TOWER, MELAPONNAGARAM 5TH STREET,
MADURAI
625016

This Contract has been prepared in two identical counterparts, one for each Party. We hereby agree to the General Terms and Conditions as detailed in the Appendix 1.

General contract agreements	
Contract start date	24-03-2019
Contract end date	23-03-2020
Contract term	1 year Contract
Invoicing	Yearly in advance
Payment Terms	Payable immediately Due net
1st-year annual price without taxes (Rs)	Rs 23,305.13
CGST @ 9%	Rs 2,097.46
SGST / UTGST @ 9%	Rs 2,097.46
Total price for first year, including applicable taxes	Rs 27,500.05
Annual Increment on Basic Price	Mutually agreed percentage on previous year's Basic Price

Signed by the Customer	Signed by KONE Elevator India Pvt Ltd
------------------------	---------------------------------------

Date	Date
Signature	Signature

Appendices: Appendix 1: Equipment Details
Appendix 2: Service Description
Appendix 3: General Conditions of Service Contract for Maintenance Agreement

For KONE Elevator India Pvt Ltd

2
KESV / P008 / A / 10

Signed For Customer



Appendix - Service Description

Scope of contract: KONE Care Standard™
Equipment type: Elevator

This contract meets all the relevant requirements of the current statutory regulations.

Description of work

KONE Modular based maintenance™	KONE Modular based maintenance™ is KONE's preventive maintenance method. Maintenance activities are done according to equipment specific maintenance plan. Pre-defined maintenance modules include the maintenance actions for each main component of the equipment.
KONE Customer Care Centre™	KONE Customer Care Center is a 24/7 helpdesk for reporting technical failures and faults in the elevators, escalators and doors or for requests of other assistance on site. KONE Customer Care Center can be easily accessed through one national phone number. KONE Customer Care Center personnel answer to service requests and dispatch KONE technicians to perform the Call-Out and Entrapment Rescue Services. Service requests for non-urgent Service Repair work is assigned to KONE field operations.
Labour for Call-outs Service	Call-out Service is designed to solve unexpected equipment failure, equipment stoppage or erratic operation, requiring immediate attention of a KONE Technician. Response Times are committed by KONE under mutual consent with customer.
Agreed Maintenance Times	Maintenance carried out during normal working hours (Monday - Saturday 08:30 - 17:30)

For KONE Elevator India Pvt Ltd.
For KONE Elevator India Pvt Ltd.

KESV / F003 / A / 10

Signed For Customer





KONE Care™

GENERAL TERMS AND CONDITIONS FOR MAINTENANCE SERVICES

1 DEFINITIONS

The following defined words and phrases shall have the following meanings:

Term	Description
"Commencement Date"	The date of commencement of the Contract
"Contract"	The contract entered into between KONE and the CUSTOMER regarding the provision of Maintenance Services
"Contract Duration Period"	The duration period of the Contract as set out in the Contract.
"Equipment"	The elevators and escalators listed in the Contract, and related components and parts that are a part of the original supply.
"Legislative Requirements"	All applicable regulatory and legislative requirements, laws, statutes, regulations and requirements and/or orders set out by any competent authority.
"Maintenance Services"	All services to be performed by KONE with respect to the Equipment pursuant to the Contract
"Normal Working Hours"	The time as specified under the "Contract Details" section
"Party" or "Parties"	The CUSTOMER and/or KONE
"Price"	Consideration payable to KONE by the CUSTOMER for the performance of the Maintenance Services

2. PROVISION OF MAINTENANCE SERVICES BY KONE

KONE shall perform the Maintenance Services as agreed to in the Contract and in these General Terms and Conditions. In performing the said services, KONE shall take all reasonable steps to maintain the Equipment in proper operating condition. KONE shall use trained and appropriately supervised personnel to perform the Maintenance Services. The Maintenance Services shall be conducted during the Normal Working Hours. If not separately agreed, any work conducted outside the Normal Working Hours is not included in the Price and shall be invoiced separately. KONE during its normal working hours, shall send at regular intervals and as frequently as the Company thinks necessary, having regard to the age, nature and condition of the equipment (but not less than 12 times per annum) a technician to systematically inspect, adjust and lubricate the parts of the equipment to the extent necessary to maintain the equipment in satisfactory working order. If not separately

agreed, any work conducted outside the Normal Working Hours is not included in the Price and shall be invoiced separately. KONE will supply all lubricants (made as per standards of KONE) necessary for this purpose

Upon notification by the customer of a breakdown or failure in the equipment, the KONE shall send, as soon as may reasonably be possible and during KONE's normal working hours, a technician to carry out necessary repairs in order to restore the equipment to satisfactory working condition.

KONE will carry out according to its standards customary annual safety test to examine all safety devices. KONE will not be required to make any other tests. KONE will neither be required to install new attachments nor to make replacements with parts of a different design to the equipment whether or not recommended or directed by Insurance Companies, or by Governmental or Non-Governmental authorities.

KONE is not expected to assume liability for injury (other than to its employees) or damage to property resulting from or caused by the equipment during its operation. KONE reserves the right to keep the control cubicle locked. The Equipment under contract will remain out of commissioning while the maintenance process is being carried out. No one will be allowed to use the Equipment during this period.

3. PROVISIONS BY THE CUSTOMER

The CUSTOMER shall promptly inform KONE of any unsatisfactory operation or performance of the Equipment, any accidents or incidents involving the Equipment or any change in the use of the Equipment. The CUSTOMER shall provide a safe and adequate working environment for KONE personnel and reasonable access to carry out the Maintenance Services. The CUSTOMER shall be responsible for all wiring in the building structure and power supply necessary for the functioning of the Equipment. The CUSTOMER shall be responsible for any power supply fluctuations or failures causing damage to the Equipment. The CUSTOMER shall comply with all applicable Legislative Requirements, including occupational safety and health regulations.

The CUSTOMER shall keep sills, machine room and pit clean. The CUSTOMER shall instruct all persons using the equipment to use it all times in accordance with KONE's reasonable instructions. The CUSTOMER shall ensure to prevent misuse or vandalism of the equipment

The CUSTOMER shall ensure that two trained persons in the building will be available for emergency rescue of trapped passengers. The CUSTOMER shall nominate two persons by name and designation for intimating breakdowns if any, to KONE with clear understanding that instructions of only such persons will be attended by KONE.

The CUSTOMER shall keep the m/c room under lock and key.

The CUSTOMER shall not to allow any other person, either his own or a third party to meddle with, repair or rectify any of the equipment components during the subsistence of this contract with the explicit understanding that any breach of this clause will relieve KONE of all further obligations under this contract.

4. PAYMENT AND ADJUSTMENT OF PRICE

For KONE Elevator India Pvt Ltd

KONE Elevator India Pvt Ltd

KESW/30087/A/10

Signed For Customer





Unless otherwise stipulated in the Contract, the Price is due annually in advance by means of cheque / DD shall be settled against submission of pro-forma invoice OR within 5 days from receipt of the invoice by the CUSTOMER. The Price is exclusive of Taxes as applicable. The Price may be adjusted annually by KONE in accordance with any increase in the cost of performing the Maintenance Services during any invoicing period. Any such variations will be made according to the price adjustment percentage agreed between the Parties. Further, KONE reserves the right to adjust the Price in the event the main purpose of use of the Equipment materially changes during the Contract Duration Period or in the event new Legislative Requirements enter into force which materially change the scope of the Maintenance Services or the costs of providing the said services. The CUSTOMER shall pay in addition to the contract price mentioned here, any tax imposed upon the CUSTOMER, or KONE or KONE's suppliers by any existing or future law, or under any statute court decision, rule or regulations becoming effective after the date of this proposal which is based upon or incident to the use, ownership or possession of the materials or equipment involved in the performance here of or service rendered hereunder.

5. DELAYED PAYMENT BY THE CUSTOMER

If the payment of any amount due under the Contract is delayed, KONE shall be entitled to charge interest on such sum at the rate of eighteen per cent (18%) per annum on amount unpaid as per payment terms after date of invoice. KONE shall also have the right, without prejudice to other remedies, to suspend immediately the provision of the Maintenance Services until the payments due to KONE (with interest) have been paid in full.

6. KONE PARTS AND COMPONENTS

All parts and components fitted by KONE will be original parts or components. If original parts or components are not available, the parts or components fitted will be of equal quality and functionality. KONE's liability to the CUSTOMER for any defects in design, materials or workmanship relating to parts and components shall be limited to the replacement of spare parts or components as defined by this Article 6.

a) If any component is rendered defective affecting equipment performance, repair or replacement of the component will be done on a chargeable basis. Any replacement which is less than Rs. 500/- will be carried out without prior sanction and a bill will be submitted. For repair or replacement value more than Rs. 500/- prior approval will be taken. Approvals must be given within 2 days of notice failing which KONE will not be responsible for any consequent breakdown or accident and such breakdowns will be attended on a chargeable basis.

Any other equipment or accessory not forming part of the initial supply of the equipment although provided as a necessary accessory by or to the customer. This includes Accessories such as EBD / KR D , Intercom , LAS ,BMS,DCS,E-Link & Group Indicators

b) It is hereby specifically agreed that KONE would not in any way be liable to replace or repair free of charge, under this contract any damage caused to all or part of the equipment as a consequence of a faulty electrical system, fire, water seepage flooding etc. In such an event all repairs and replacements as may be necessitated would be carried out at the cost and expense of the customer. It is recommended that the customer should take adequate protection from insurance or similar companies to safe guard the equipment for damages that would occur due to such causes. In such an event the cost of repair or replacement should be reimbursed to KONE without any conditions or limitations

c) In the event KONE consider themselves unable to supply any materials or parts due to obsolescence or if they have been permanently taken out of production by the original supplier, then this agreement shall forthwith terminate without prejudice to KONE's accrued rights and without any liability to KONE for such termination.

7. LIMITATION OF LIABILITY

Notwithstanding any other provisions or indemnities in this Contract, in no event shall KONE be liable to the other party for any loss of profit, use, contracts, business, customers, good will, contractual liabilities of others or for any indirect or consequential loss or damage, which may be suffered by the other party in connection with the Contract. KONE's maximum aggregate liability under or in relation with this Contract shall in no event exceed an amount equal to one year's Contract value per equipment.

8. FORCE MAJEURE AND RELEASE FROM RESPONSIBILITY

KONE shall not be liable for any failure to fulfill any of its obligations under the Contract to the extent that such fulfillment is prevented by circumstances beyond KONE's reasonable control, including but not limited to acts of God, epidemic, acts of government, war, civil commotion, terrorism, material shortages, transportation delays, labour unrest, theft, vandalism, misuse of Equipment, failure of incoming power supply, fire, flood, adverse climate conditions or natural disasters.

9. PROPERTY RIGHTS

The proprietary rights to any drawings, technical documentation, software or other intellectual property provided by KONE in the course of and in connection with performance of the Maintenance Services, shall remain solely with KONE. KONE is not expected to assume possession or Management of any part of the equipment and the customer remains exclusively as the owner.

10. TERMINATION OF CONTRACT

The Contract shall remain in force for the Contract Duration Period, unless cancelled in writing by either party ninety (90) days prior to the desired date of termination. Either Party may terminate the Contract, without satisfying the above time requirements, by giving a written notice to the other Party in the event that the other Party goes into liquidation either compulsorily or voluntarily, or a receiver, administrator or administrative receiver is appointed in respect of the whole or any part of its assets, or if the other Party commits a material breach of the Contract and the said breach has not been remedied within thirty (30) days after receipt of written notice setting forth particulars describing the alleged breach. Further, KONE has the right to terminate the Contract in the event the main purpose of use of the Equipment has materially changed during the Contract Duration Period or the Equipment is serviced or repaired by a third party without the prior written approval by KONE during the Contract Duration Period or the ownership of the building where the Equipment is located is changed.

In the event that the Contract is terminated by either Party as set out above, the CUSTOMER's obligations to make payments due under the Contract shall survive the termination and KONE shall be entitled to receive payments from the CUSTOMER for any Maintenance Services performed before the effective date of termination. KONE, in turn, shall return any payments made by the CUSTOMER for Maintenance Services not yet performed, except in the event





of termination due to the CUSTOMER's default. In the event the Contract is terminated for any reason whatsoever any Equipment specific maintenance productivity tools will be removed, unless the Customer wishes to purchase such tools at the prevailing market price.

In all circumstances where the Contract is terminated other than for KONE's breach, insolvency or repudiation, a termination fee shall become payable by the CUSTOMER, without prejudice to any of KONE's other rights. The said termination fee is equal to 30% of the Price payable by the Customer for the remaining term of the Contract but for the termination. The Parties agree that the said termination fee is a reasonable pre-estimate of the loss suffered by KONE as a result of the termination.

11. APPLICABLE LAW

In the event of difference or dispute arising out of, under or in connection with this contract / agreement, over the right of obligation of parties hereto, the dispute or difference shall be referred to the Arbitration of a Sole Arbitrator, to be appointed by KONE. The Contract will be governed by the laws of India and the courts of Chennai shall have sole jurisdiction over any disputes between the Parties relating to the Contract, and the Provisions of the Arbitration & Conciliation Act 1996 shall be applicable to such Arbitration.

12. RESCUE TRAINING

As a part of maintenance, KONE intends to provide basic training to CUSTOMER with respect to rescuing passengers entrapped in elevators under certain limited circumstances.

The objective of providing rescue training is to enable the CUSTOMER to understand and identify the risks involved in relation to elevator entrapment situations and related rescue operations, what kind of rescue operation is needed, whether the needed rescue operation can be conducted by the CUSTOMER, and further provide knowledge and teach practical skills needed in the rescue operations. The training would include on site demonstration of limited rescue operations that needs to be taken into account when rescuing passengers including the "dos and don'ts".

This training shall be provided by KONE to the CUSTOMER, as a one-time exercise free of cost at the request of CUSTOMER, for any two of the representatives nominated by the CUSTOMER. Subsequent training required to be provided by KONE at any future point of time shall be chargeable on a mutually agreeable basis.

After providing training, KONE and CUSTOMER shall record the fact of having provided such basic training in the format prescribed by KONE. This shall include the names and positions of the CUSTOMER's representative who have participated in the said training. KONE shall always keep the determining record of who have participated in the training. The CUSTOMER or their representative who has been so trained can thereafter engage in rescuing passengers entrapped in an elevator using solely the methods as trained by KONE.

The CUSTOMER and their representatives acknowledge and understand that they shall not engage in rescue operations for which they are not trained and shall not attempt to rescue any persons following processes, other than those for which they have been trained by KONE. This training shall be valid only for the validity of this contract.

CUSTOMER hereby voluntarily release, forever discharge and agree to indemnify and hold harmless KONE, its directors, officers, employees, agents, subcontractors, volunteers and all other persons or entities acting in any capacity on behalf of KONE from any and all liability, claims, demands or causes of action which may be in any way connected with the participation of the CUSTOMER and/or their representatives in the training activity including all such claims which allege negligent acts or omissions of KONE.

It is specifically agreed that the training provided by KONE shall not absolve the CUSTOMER or their representatives from any negligent and/or any acts of omission or commission that may result in any accident / cause damage either to the entrapped passengers or to the property. KONE shall not be held responsible for any consequences arising out of rescue undertaken by the CUSTOMER or their representatives whether the rescue is happening before, during or after any training provided by KONE. CUSTOMER agrees to indemnify KONE, its directors, officers, employees, agents, subcontractors, volunteers and all other persons or entities acting in any capacity on behalf of KONE against claims, demand, prosecution and/or any charge arising therefrom.

13. MISCELLANEOUS

The Contract constitutes the entire agreement between the Parties, and supersedes all prior negotiations, understandings, representations, and agreements between the Parties, if any. The CUSTOMER represents and warrants that in deciding to enter into the Contract, the CUSTOMER has not relied on any information supplied or statements made by KONE except those set forth in the Contract. The Contract may be amended or varied only by a written instrument signed by duly authorized representatives of both Parties. Any purchase order issued by the CUSTOMER in connection with the Maintenance Services shall be deemed to be issued for the CUSTOMER's administrative billing purposes only, and the Parties hereby intend that the terms and conditions of the Contract shall exclusively govern any services to be provided hereunder. None of the conditions of the Contract shall be considered waived by either Party unless such waiver is given in writing by the Party. No such waiver shall be a waiver of any past or future default, breach or modification of any of the conditions of the Contract, unless expressly stipulated in such waiver. This Contract can be freely assigned by KONE to any other company within the KONE group without the prior consent of the Customer. Notwithstanding any transfer of ownership of the building where the Equipment is located or change of the property manager of the said building, this Contract will continue in full force and effect until the end of the Contract Duration Period.





4

cc

SARADA KRISHNA HOMOEOPATHIC
MEDICAL COLLEGE
KULASEKHARAM
KANYAKUMARI
629161
India

KONE Elevator India Pvt Ltd

NO.1,PANDIAN TOWER,
MELAPONNAGARAM 5TH
STREET,
MADURAI
625016
Tel:0452 2361617
Contact person:S K Kumaresh
Contact person mob:+91 80560
41261

Contract No: 41424123

Quotation No. T-0002686872 v1

08.12.2018

KONE Care™ contract

Dear

We wish to express our sincere gratitude to you for continuing to be KONE customer.

Your continued patronage has helped us to provide uninterrupted services to KONE Elevators / Escalators in your premises. It has been our endeavor to ensure a safe and comfortable ride for all users of KONE Equipment in your building.

It is quite some time, since we signed the full set of contract documents; hence it is time for us to renew our agreement towards the scope, terms & conditions of the referred Maintenance Contract. Further, the costs of labour and material have seen a substantial upward movement during the last one year. We are forced to pass on a certain portion of the increased cost to continue maintaining our service routines. Accordingly, we propose a nominal increase over the present contract value.

Enclosed please find the contract document. We would request you to please sign the copies on each page and return to us to enable us, process your instruction as speedily and efficiently as possible.

However should you have any questions regarding our proposal, or require any further information, please do not hesitate to contact us.

Yours sincerely
KONE Elevator India Pvt Ltd
S K Kumaresh
Engineer - Service
sk.kumaresh@kone.com

For KONE Elevator India Pvt Ltd

KESV / F006 / A / 18

Signed For Customer



KONE Elevator India Pvt Ltd



KONE Care Standard™

Sold to
 SARADA KRISHNA
 HOMOEOPATHIC
 KULASEKHARAM
 KANYAKUMARI
 TAMIL NADU
 629161
 India

Invoice to
 SARADA KRISHNA
 HOMOEOPATHIC MEDICAL
 COLLEGE
 KULASEKHARAM
 629161
 KANYAKUMARI

and contractor:

KONE Elevator India Pvt Ltd
 NO 1,PANDIAN TOWER, MELAPONNAGARAM 5TH STREET,
 MADURAI
 625016

This Contract has been prepared in two identical counterparts, one for each Party. We hereby agree to the General Terms and Conditions as detailed in the Appendix 1.

General contract agreements	
Contract start date	26-12-2018
Contract end date	25-12-2019
Contract term	1 year Contract
Invoicing	Yearly in advance
Payment Terms	Payable immediately Due net
1st-year annual price without taxes (Rs)	Rs 28,105.97
CGST @ 9%	Rs 2,529.54
SGST / UTGST @ 9%	Rs 2,529.54
Total price for first year, including applicable taxes	Rs 33,165.04
Annual increment on Basic Price	Mutually agreed percentage on previous year's Basic Price

Signed by the Customer	Signed by KONE Elevator India Pvt Ltd
------------------------	---------------------------------------

Date	Date
Signature	Signature

- Appendices:
- Appendix 1: Equipment Details
 - Appendix 2: Service Description
 - Appendix 3: General Conditions of Service Contract for Maintenance Agreement

For KONE Elevator India Pvt Ltd

2
 KESV / F008 / A / 10

Signed For Customer





Asset List - Equipment Details

The following equipment shall be covered by this contract

Scope of contract: KONE Care Standard™
Equipment type: Elevator

Equipment number	Address	capacity (Kg)	Number of landing doors	1st-year annual price without taxes (Rs)
42461672	KULASEKHARAM, KANYAKUMARI 629161	884	5	Rs 28,105.97

For KONE Elevator India Pvt Ltd

3
KESV / F008 / A / 10

Signed For Customer

For KONE Elevator India Pvt. Ltd.
KONE ID




Appendix - Service Description

Scope of contract: KONE Care Standard™
Equipment type: Elevator

This contract meets all the relevant requirements of the current statutory regulations.

Description of work

KONE Modular based maintenance™	KONE Modular based maintenance™ is KONE's preventive maintenance method. Maintenance activities are done according to equipment specific maintenance plan. Pre-defined maintenance modules include the maintenance actions for each main component of the equipment.
KONE Customer Care Centre™	KONE Customer Care Center is a 24/7 helpdesk for reporting technical failures and faults in the elevators, escalators and doors or for requests of other assistance on site. KONE Customer Care Center can be easily accessed through one national phone number.
Labour for Call-outs Service	KONE Customer Care Center personnel answer to service requests and dispatch KONE technicians to perform the Call-Out and Entrapment Rescue Services. Service requests for non-urgent Service Repair work is assigned to KONE field operations. Call-out Service is designed to solve unexpected equipment failure, equipment stoppage or erratic operation, requiring immediate attention of a KONE Technician. Response Times are committed by KONE under mutual consent with customer.
Agreed Maintenance Times	Maintenance carried out during normal working hours (Monday - Saturday 08:30 - 17:30)





KONE Care™

GENERAL TERMS AND CONDITIONS FOR MAINTENANCE SERVICES

1 DEFINITIONS

The following defined words and phrases shall have the following meanings:

Term	Description
"Commencement Date"	The date of commencement of the Contract
"Contract"	The contract entered into between KONE and the CUSTOMER regarding the provision of Maintenance Services
"Contract Duration Period"	The duration period of the Contract as set out in the Contract.
"Equipment"	The elevators and escalators listed in the Contract, and related components and parts that are a part of the original supply.
"Legislative Requirements"	All applicable regulatory and legislative requirements, laws, statutes, regulations and requirements and/or orders set out by any competent authority.
"Maintenance Services"	All services to be performed by KONE with respect to the Equipment pursuant to the Contract
"Normal Working Hours"	The time as specified under the "Contract Details" section
"Party" or "Parties"	The CUSTOMER and/or KONE
"Price"	Consideration payable to KONE by the CUSTOMER for the performance of the Maintenance Services

2. PROVISION OF MAINTENANCE SERVICES BY KONE

KONE shall perform the Maintenance Services as agreed to in the Contract and in these General Terms and Conditions. In performing the said services, KONE shall take all reasonable steps to maintain the Equipment in proper operating condition. KONE shall use trained and appropriately supervised personnel to perform the Maintenance Services. The Maintenance Services shall be conducted during the Normal Working Hours. If not separately agreed, any work conducted outside the Normal Working Hours is not included in the Price and shall be invoiced separately. KONE during its normal working hours, shall send at regular intervals and as frequently as the Company thinks necessary, having regard to the age, nature and condition of the equipment (but not less than 12 times per annum), a technician to systematically inspect, adjust and lubricate the parts of the equipment to the extent necessary to maintain the equipment in satisfactory working order. If not separately

agreed, any work conducted outside the Normal Working Hours is not included in the Price and shall be invoiced separately. KONE will supply all lubricants (made as per standards of KONE) necessary for this purpose

Upon notification by the customer of a breakdown or failure in the equipment, the KONE shall send, as soon as may reasonably be possible and during KONE's normal working hours, a technician to carry out necessary repairs in order to restore the equipment to satisfactory working condition.

KONE will carry out according to its standards customary annual safety test to examine all safety devices. KONE will not be required to make any other tests. KONE will neither be required to install new attachments nor to make replacements with parts of a different design to the equipment whether or not recommended or directed by Insurance Companies, or by Governmental or Non-Governmental authorities.

KONE is not expected to assume liability for injury (other than to its employees) or damage to property resulting from or caused by the equipment during its operation. KONE reserves the right to keep the control cubicle locked. The Equipment under contract will remain out of commissioning while the maintenance process is being carried out. No one will be allowed to use the Equipment during this period.

3. PROVISIONS BY THE CUSTOMER

The CUSTOMER shall promptly inform KONE of any unsatisfactory operation or performance of the Equipment, any accidents or incidents involving the Equipment or any change in the use of the Equipment. The CUSTOMER shall provide a safe and adequate working environment for KONE personnel and reasonable access to carry out the Maintenance Services. The CUSTOMER shall be responsible for all wiring in the building structure and power supply necessary for the functioning of the Equipment. The CUSTOMER shall be responsible for any power supply fluctuations or failures causing damage to the Equipment. The CUSTOMER shall comply with all applicable Legislative Requirements, including occupational safety and health regulations.

The CUSTOMER shall keep sills, machine room and pit clean. The CUSTOMER shall instruct all persons using the equipment to use it all times in accordance with KONE's reasonable instructions. The CUSTOMER shall ensure to prevent misuse or vandalism of the equipment.

The CUSTOMER shall ensure that two trained persons in the building will be available for emergency rescue of trapped passengers. The CUSTOMER shall nominate two persons by name and designation for intimating breakdowns if any, to KONE with clear understanding that instructions of only such persons will be attended by KONE.

The CUSTOMER shall keep the m/c room under lock and key.

The CUSTOMER shall not to allow any other person, either his own or a third party to meddle with, repair or rectify any of the equipment components during the subsistence of this contract with the explicit understanding that any breach of this clause will relieve KONE of all further obligations under this contract.

4. PAYMENT AND ADJUSTMENT OF PRICE



Unless otherwise stipulated in the Contract, the Price is due annually in advance by means of cheque / DD shall be settled against submission of pro-forma invoice OR within 5 days from receipt of the invoice by the CUSTOMER. The Price is exclusive of Taxes as applicable. The Price may be adjusted annually by KONE in accordance with any increase in the cost of performing the Maintenance Services during any invoicing period. Any such variations will be made according to the price adjustment percentage agreed between the Parties. Further, KONE reserves the right to adjust the Price in the event the main purpose of use of the Equipment materially changes during the Contract Duration Period or in the event new Legislative Requirements enter into force which materially change the scope of the Maintenance Services or the costs of providing the said services. The CUSTOMER shall pay in addition to the contract price mentioned here, any tax imposed upon the CUSTOMER, or KONE or KONE's suppliers by any existing or future law, or under any statute court decision, rule or regulations becoming effective after the date of this proposal which is based upon or incident to the use, ownership or possession of the materials or equipment involved in the performance here of or service rendered hereunder.

5. DELAYED PAYMENT BY THE CUSTOMER

If the payment of any amount due under the Contract is delayed, KONE shall be entitled to charge interest on such sum at the rate of eighteen per cent (18%) per annum on amount unpaid as per payment terms after date of invoice. KONE shall also have the right, without prejudice to other remedies, to suspend immediately the provision of the Maintenance Services until the payments due to KONE (with interest) have been paid in full.

6. KONE PARTS AND COMPONENTS

All parts and components fitted by KONE will be original parts or components. If original parts or components are not available, the parts or components fitted will be of equal quality and functionality. KONE's liability to the CUSTOMER for any defects in design, materials or workmanship relating to parts and components shall be limited to the replacement of spare parts or components as defined by this Article 6.

a) If any component is rendered defective affecting equipment performance, repair or replacement of the component will be done on a chargeable basis. Any replacement which is less than Rs. 500/- will be carried out without prior sanction and a bill will be submitted. For repair or replacement value more than Rs. 500/- prior approval will be taken. Approvals must be given within 2 days of notice failing which KONE will not be responsible for any consequent breakdown or accident and such breakdowns will be attended on a chargeable basis.

Any other equipment or accessory not forming part of the initial supply of the equipment although provided as a necessary accessory by or to the customer. This includes Accessories such as EBD / KRD , Intercom , LAS BMS,DCS,E-Link & Group Indicators

b) It is hereby specifically agreed that KONE would not in any way be liable to replace or repair free of charge, under this contract any damage caused to all or part of the equipment as a consequence of a faulty electrical system, fire, water seepage flooding etc. In such an event all repairs and replacements as may be necessitated would be carried out at the cost and expense of the customer. It is recommended that the customer should take adequate protection from insurance or similar companies to safe guard the equipment for damages that would occur due to such causes. In such an event the cost of repair or replacement should be reimbursed to KONE without any conditions or limitations

c) In the event KONE consider themselves unable to supply any materials or parts due to obsolescence or if they have been permanently taken out of production by the original supplier then this agreement shall forthwith terminate without prejudice to KONE's accrued rights and without any liability to KONE for such termination.

7. LIMITATION OF LIABILITY

Notwithstanding any other provisions or indemnities in this Contract, in no event shall KONE be liable to the other party for any loss of profit, use, contracts, business, customers, good will, contractual liabilities of others or for any indirect or consequential loss or damage, which may be suffered by the other party in connection with the Contract. KONE's maximum aggregate liability under or in relation with this Contract shall in no event exceed an amount equal to one year's Contract value per equipment.

8. FORCE MAJEURE AND RELEASE FROM RESPONSIBILITY

KONE shall not be liable for any failure to fulfill any of its obligations under the Contract to the extent that such fulfillment is prevented by circumstances beyond KONE's reasonable control, including but not limited to acts of God, epidemic, acts of government, war, civil commotion, terrorism, material shortages, transportation delays, labour unrest, theft, vandalism, misuse of Equipment, failure of incoming power supply, fire, flood, adverse climate conditions or natural disasters.

9. PROPERTY RIGHTS

The proprietary rights to any drawings, technical documentation, software or other intellectual property provided by KONE in the course of and in connection with performance of the Maintenance Services, shall remain solely with KONE. KONE is not expected to assume possession or Management of any part of the equipment and the customer remains exclusively as the owner.

10. TERMINATION OF CONTRACT

The Contract shall remain in force for the Contract Duration Period, unless cancelled in writing by either party ninety (90) days prior to the desired date of termination. Either Party may terminate the Contract, without satisfying the above time requirements, by giving a written notice to the other Party in the event that the other Party goes into liquidation either compulsorily or voluntarily, or a receiver, administrator or administrative receiver is appointed in respect of the whole or any part of its assets, or if the other Party commits a material breach of the Contract and the said breach has not been remedied within thirty (30) days after receipt of written notice setting forth particulars describing the alleged breach. Further, KONE has the right to terminate the Contract in the event the main purpose of use of the Equipment has materially changed during the Contract Duration Period or the Equipment is serviced or repaired by a third party without the prior written approval by KONE during the Contract Duration Period or the ownership of the building where the Equipment is located is changed.

In the event that the Contract is terminated by either Party as set out above, the CUSTOMER's obligations to make payments due under the Contract shall survive the termination and KONE shall be entitled to receive payments from the CUSTOMER for any Maintenance Services performed before the effective date of termination. KONE in turn, shall return any payments made by the CUSTOMER for Maintenance Services not yet performed, except in the event



of termination due to the CUSTOMER's default. In the event the Contract is terminated for any reason whatsoever any Equipment specific maintenance productivity tools will be removed unless the Customer wishes to purchase such tools at the prevailing market price.

In all circumstances where the Contract is terminated other than for KONE's breach, insolvency or repudiation, a termination fee shall become payable by the CUSTOMER, without prejudice to any of KONE's other rights. The said termination fee is equal to 30% of the Price payable by the Customer for the remaining term of the Contract but for the termination. The Parties agree that the said termination fee is a reasonable pre-estimate of the loss suffered by KONE as a result of the termination.

11. APPLICABLE LAW

In the event of difference or dispute arising out of, under or in connection with this contract / agreement, over the right of obligation of parties hereto, the dispute or difference shall be referred to the Arbitration of a Sole Arbitrator, to be appointed by KONE. The Contract will be governed by the laws of India and the courts of Chennai shall have sole jurisdiction over any disputes between the Parties relating to the Contract, and the Provisions of the Arbitration & Conciliation Act 1996 shall be applicable to such Arbitration.

12. RESCUE TRAINING

As a part of maintenance, KONE intends to provide basic training to CUSTOMER with respect to rescuing passengers entrapped in elevators under certain limited circumstances. The objective of providing rescue training is to enable the CUSTOMER to understand and identify the risks involved in relation to elevator entrapment situations and related rescue operations, what kind of rescue operation is needed, whether the needed rescue operation can be conducted by the CUSTOMER, and further provide knowledge and teach practical skills needed in the rescue operations. The training would include on site demonstration of limited rescue operations that needs to be taken into account when rescuing passengers including the "dos and don'ts". This training shall be provided by KONE to the CUSTOMER as a one-time exercise free of cost at the request of CUSTOMER, for any two of the representatives nominated by the CUSTOMER. Subsequent training required to be provided by KONE at any future point of time shall be chargeable on a mutually agreeable basis. After providing training, KONE and CUSTOMER shall record the fact of having provided such basic training in the format prescribed by KONE. This shall include the names and positions of the CUSTOMER's representative who have participated in the said training. KONE shall always keep the determining record of who have participated in the training. The CUSTOMER or their representative who has been so trained can thereafter engage in rescuing passengers entrapped in an elevator using solely the methods as trained by KONE.

The CUSTOMER and their representatives acknowledge and understand that they shall not engage in rescue operations for which they are not trained and shall not attempt to rescue any persons following processes, other than those for which they have been trained by KONE. This training shall be valid only for the validity of this contract.

CUSTOMER hereby voluntarily release, forever discharge and agree to indemnify and hold harmless KONE, its directors, officers, employees, agents, subcontractors, volunteers and all other persons or entities acting in any capacity on behalf of KONE from any and all liability, claims, demands or causes of action which may be in any way connected with the participation of the CUSTOMER and/or their representatives in the training activity including all such claims which allege negligent acts or omissions of KONE. It is specifically agreed that the training provided by KONE shall not absolve the CUSTOMER or their representatives from any negligent and/or any acts of omission or commission that may result in any accident / cause damage either to the entrapped passengers or to the property. KONE shall not be held responsible for any consequences arising out of rescue undertaken by the CUSTOMER or their representatives whether the rescue is happening before, during or after any training provided by KONE. CUSTOMER agrees to indemnify KONE, its directors, officers, employees, agents, subcontractors, volunteers and all other persons or entities acting in any capacity on behalf of KONE against claims, demand, prosecution and/or any charge arising therefrom.

13. MISCELLANEOUS

The Contract constitutes the entire agreement between the Parties, and supersedes all prior negotiations, understandings, representations, and agreements between the Parties, if any. The CUSTOMER represents and warrants that in deciding to enter into the Contract, the CUSTOMER has not relied on any information supplied or statements made by KONE except those set forth in the Contract. The Contract may be amended or varied only by a written instrument signed by duly authorised representatives of both Parties. Any purchase order issued by the CUSTOMER in connection with the Maintenance Services shall be deemed to be issued for the CUSTOMER's administrative billing purposes only and the Parties hereby intend that the terms and conditions of the Contract shall exclusively govern any services to be provided hereunder. None of the conditions of the Contract shall be considered waived by either Party unless such waiver is given in writing by the Party. No such waiver shall be a waiver of any past or future default, breach or modification of any of the conditions of the Contract, unless expressly stipulated in such waiver. This Contract can be freely assigned by KONE to any other company within the KONE group without the prior consent of the Customer. Notwithstanding any transfer of ownership of the building where the Equipment is located or change of the property manager of the said building, this Contract will continue in full force and effect until the end of the Contract Duration Period.

For KONE Elevator India Pvt. Ltd.

S. K. Kumarash
For KONE Elevator India Pvt. Ltd.

7
KERN / 2009 / A / 10

Signed For Customer





JOE ASSOCIATES

CONSULTING ARCHITECTS, ENGINEERS & VALUERS
BERACHAH TOWER, CHIRAYANKUZZHI JN., KANJIRACODE P.O, PIN : 629 155.

Ph : (O) 04651 - 271080, (R) 273060 Mob : 9443607580, Email : pisonjoe@yahoo.com.

Website : www.joeassociates.com

Er. J. PISON JOE, B.Sc., B.E., M.I.E., C.Eng (I), F.I.V.


Govt. Regd. Engineer | Govt. Regd. Valuer | Indian Bank Panel Valuer | Karnataka Bank Panel Valuer | Federal Bank Panel Valuer |
Chartered Engineer | Approved Valuer | H.D.F.C. Panel Valuer | Corporation Bank Panel Valuer | S.B.I. Panel Valuer | L.I.C. Panel Valuer |

STATEMENT ACCOMPANYING FORM - A

S2/C/18


1. Door Numbers of Building : 3 - 101, 3 - 101 /2 & 3
2. Location (street no., ward no.) Name of place i.e Corp./ Municipality / Panchayat Union / Township : Sarada Krishna Homeopathic Medical College,
Kulasekharam, R.S. Nos: 392 / 2 A & 2B2
3. Year of Construction : 2000
4. Approximate cost of construction as per current rates : ₹ 2,10,00,000/-
5. Purpose for which the building is being used or proposed to be used. : College.
6. Purpose for which the building has hitherto been Used. : College.
7. Details of construction of the building (particulars of the foundation soil, specification of the various parts under) :
 - i. Foundation : R.C.C. Column Footing
 - ii. Basement : R.R.M. in Cement Mortar
 - iii. Superstructure : R.C.C. Framed Structure & C.B.. in C.M.
 - iv. Floor : Partly Ceramic Tiles & Cement
 - v. Roof : R.C.C.
8. Period for which license is required : 01.04.2018 to 31.03.2021 (Three Years)
9. Number of persons to be accommodated : 700 Nos.

Date: 05.03.2018


Govt. Registered Engineer

Certified that no addition or alteration has been made to the existing building or portion of the building for which this application has been made. I undertake to obtain fresh license, in case addition or alteration is made to the existing building or portion of the building. I also undertake to obtain fresh license, if the purpose for which the license was granted for the use of the building originally is changed or altered subsequently.

Date: 05.03.2018


Signature of the owner of the building


FORM - C

STATEMENT TO ACCOMPANY THE CERTIFICATE OF STRUCTURAL SOUNDNESS

Location of the Building Door No., Street and Place.	Age of the building	No. of stories	Short description	Certified purpose the building stability	Certified period for the building stability	No. of persons to be accommodated	Remarks
3 -101, 3-101/2 & 3 Sarada Krishna Homeopathic Medical College, Kulasekharam, R.S.Nos: 392/2 A & 2B2	17 Years	G.F., F.F. S.F. & T.F. Four stories	R.C.C. framed structure, partly Country wood & Steel doors, windows R.C.C. roof	College	01.04.2018 to 31.03.2021 (Three Years)	700 Nos.	-----

DS/23/18

Date: 05.03.2018


Govt. Registered Engineer

Engineer
No. 1
1-3-4
01/03/2018



JOE ASSOCIATES

CONSULTING ARCHITECTS, ENGINEERS & VALUERS
BERACHAH TOWER, CHIRAYANKUZHI JN., KANJIRACODE P.O, PIN : 629 155.

Ph : (0) 04651 - 271080, (R) 273080 Mob : 9443607580, Email : pisonjoe@yahoo.com.

Website : www.joeassociates.com

Er. J. PISON JOE, B.Sc., B.E., M.I.E., C.Eng (I), F.I.V.

Govt. Regd. Engineer | Govt. Regd. Valuer | Indian Bank Panel Valuer | Karnataka Bank Panel Valuer | Federal Bank Panel Valuer |
Chartered Engineer | Approved Valuer | H.D.F.C. Panel Valuer | Corporation Bank Panel Valuer | S.B.I. Panel Valuer | L.I.C. Panel Valuer

STATEMENT ACCOMPANYING FORM - A

S1/C/18

1. Door Number of Building : 3 - 101/ 5 & 6
2. Location (street no., ward no.) Name of place i.e Corp./ Municipality / Panchayat Union / Township : Sarada Krishna Homeopathic Medical College Hospital,
Kulasekharam, R.S. Nos: 392 / 2A & 2B
3. Year of Construction : 2006
4. Approximate cost of construction as per current rates : ₹ 1,80,00,000/-
5. Purpose for which the building is being used or proposed to be used. : Hospital.
6. Purpose for which the building has hitherto been Used. : Hospital.
7. Details of construction of the building (particulars of the foundation soil, specification of the various parts under)
 - i. Foundation : R.C.C. Column Footing.
 - ii. Basement : R.R.M. in Cement Mortar.
 - iii. Superstructure : R.C.C. Framed Structure & C.B.. in C.M.
 - iv. Floor : Partly Ceramic Tiles & Cement
 - v. Roof : R.C.C.
8. Period for which license is required : 01.04.2018 to 31.03.2021 (Three Years)
9. Number of persons to be accommodated : 500 Nos.


Date: 05.03.2018


Govt. Registered Engineer

Er. PISON JOE
Govt. Registered Engineer
CONSULTING ARCHITECTS, ENGINEERS & VALUERS

Certified that no addition or alteration has been made to the existing building or portion of the building for which this application has been made. I undertake to obtain fresh license, in case addition or alteration is made to the existing building or portion of the building. I also undertake to obtain fresh license, if the purpose for which the license was granted for the use of the building originally is changed or altered subsequently.

Date: 05.03.2018


Signature of the owner of the building

FORM - B

Form of certificate of Structural Soundness under sub-section 4 of the Madras Public Buildings (Licensing) Act 1965, (Madras Act 13 of 1965) referred to in Rule of the Madras Public Building (Licensing) Rules 1966.

I certify that I have inspected the buildings mentioned in the statement and furnish below its technical features:

1. FOUNDATION

- i. Depth below Ground Level : 2.10m
- ii. Nature of soil met with at foundation level : Hard Gravelly Soil
- iii. Pressure of foundation : 36.82 t/m²
- iv. Probable S.B Capacity of the strata met with at foundation : 45 t/m²

2. SUPERSTRUCTURE

- i. Maximum stress in the most critical section of masonry and Concrete : Does not arise since it is a framed structure
- ii. Nature of masonry and Concrete and its safe permissible stress : -----

The structural soundness of the building has been verified by me with reference to ISI loading standard 875 (Latest version) and other relevant INDIAN STANDARDS CODE OF PRACTICE and I declare the building is structurally sound to be used as Public Building for the purpose noted in the statement. A license may be granted for the period from **01.04.2018 to 31.03.2021** (Three Years)

05/03/18

Date: 05.03.2018



Govt. Registered Engineer

ER. P. S. SIVAN
Govt. Reg.
105
05/03/2018

FORM - C

STATEMENT TO ACCOMPANY THE CERTIFICATE OF STRUCTURAL SOUNDNESS

Location of the Building Door No., Street and Place.	Age of the building	No. of stories	Short description	Certified purpose the building stability	Certified period for the building stability	No. of persons to be accommodated	Remarks
3 - 101/ 5 & 6 Sarada Krishna Homeopathic Medical College Hospital, Kulasekharam, R.S.Nos: 392/2 A & 2B2	11 Years	B.F. G.F. F.F. & S.F. Four stories	R.C.C. framed structure, partly Country wood & Steel doors, windows R.C.C. roof	Hospital	01.04.2018 to 31.03.2021 (Three Years)	500 Nos.	—

05/03/18

Date: 05.03.2018


Govt. Registered Engineer

EM...
0 11 5
...
CH...
...



JOE ASSOCIATES

CONSULTING ARCHITECTS, ENGINEERS & VALUERS

BERACHAH TOWER, CHIRAYANKUZHI JN., KANJIRACODE P.O, PIN : 629 155.

Ph : (0) 04651 - 271060, (R) 273080 Mob : 9443607580, Email : pisonjoe@yahoo.com.

Er. J. PISON JOE, B.Sc., B.E., M.I.E., C.Eng (I), F.I.V.

Website : www.joeassociates.com

Govt. Regd. Engineer | Govt. Regd. Valuer | Indian Bank Panel Valuer | Karnataka Bank Panel Valuer | Federal Bank Panel Valuer |
Chartered Engineer | Approved Valuer | H.D.F.C. Panel Valuer | Corporation Bank Panel Valuer | S.B.I. Panel Valuer | L.I.C. Panel Valuer

STATEMENT ACCOMPANYING FORM - A

S20/C/17

1. Door Number of Building : 3 - 101 / 1
2. Location (street no. ward no.) Name of place i.e. Corp./ Municipality/ Panchayat Union / Township : Sarada Krishna Homeopathic Medical College Gents Hostel,
Kulasekharam & P.O. R.S. Nos. 393 / 7
3. Year of Construction : G.F.; F.F.; S.F.;-2008, T.F.;- 2017
4. Approximate cost of construction as per current rates : Rs. 67,50,000/-
5. Purpose for which the building is being used or proposed to be used. : Gents Hostel
6. Purpose for which the building has hitherto been used : Gents Hostel
7. Details of construction of the building (particulars of the foundation soil, specification of the various parts under):
 - i. Foundation : R.C.C. Column Footing & R.R.M in C. M.
 - ii. Basement : R.C.C. Column & R.R.M. in C.M.
 - iii. Superstructure : R.C.C. Column & B.W. in C.M.
 - iv. Floor : Marble & Marbonite Flooring
 - v. Roof : R.C.C.
8. Period for which license is required : 01.08.2017 to 31.07.2020 (Three Years)
9. Number of persons to be accommodated : 100 Nos.

Date: 17.07.2017

Govt. Registered Engineer

Er. J. PISON JOE, B.Sc., B.E., M.I.E., C.Eng (I), F.I.V.

Govt. Registered Valuer

Ind. Bank Panel Valuer

CHIRAYANKUZHI, MATHANDAM Ph. 04651-271060

Certified that no addition or alteration has been made to the existing building or portion of the building for which this application has been made. I undertake to obtain fresh license, in case addition or alteration is made to the existing building or portion of the building. I also undertake to obtain fresh license, if the purpose for which the license was granted for the use of the building originally is changed or altered subsequently.

Date: 17-07-17

Signature of the owner of the building

FORM - B

Form of certificate of Structural Soundness under sub-section 4 of the Madras Public Buildings (Licensing) Act 1965, (Madras Act 13 of 1965) referred to in Rule of the Madras Public Building Rules 1966.

I certify that I have inspected the buildings mentioned in the statement and furnish below its technical features:

1. FOUNDATION

- i. Depth below Ground Level : 1.95 m
- ii. Nature of soil met with at foundation level : Hard Gravelly Soil
- iii. Pressure of foundation : 33.97 t/m²
- iv. Probable S.B Capacity of the strata met with at foundation : 45 t/m²

2. SUPERSTRUCTURE

- i. Maximum stress in the most critical section of masonry and Concrete : Does not arise since it is a framed structure
- ii. Nature of masonry and Concrete and its safe permissible stress : -----

The structural soundness of the building has been verified by me with reference to ISI loading Standard 875 (Latest version) and other relevant INDIAN STANDARDS CODE OF PRACTICE and I declare the building is structurally sound to be used as Public Building for the purpose noted in the statement. A license may be granted for the period from **01.08.2017 to 31.07.2020** (Three Years)

17/07/17

Date: 17.07.2017



Govt. Registered Engineer

Er. J. PISON JOE, B.TECH., M.A.S.T., M.C.E., I.E.I.V.

Govt. Registered Engineer

JOE ASSOCIATES

CHIRAYANKUZZHI, MARTHANDAM Ph. 04651-271080.

FORM - C

STATEMENT TO ACCOMPANY THE CERTIFICATE OF STRUCTURAL SOUNDNESS

Location of the Building Door No., Street and Place.	Age of the building	No. of stories	Short description	Certified purpose the building stability	Certified period for the building stability	No. of persons to be accommodated	Remarks
D. No. 3 - 101 / 1 Sarada Krishna Homeopathic Medical College, Gents Hostel, Kulasekharam & P.O. R.S. No. 393 / 7	8 Years	G.F., F.F., S.F., & T.F.: Four stories	R.C.C. Framed Structure, Country wood doors, windows,	Gents Hostel	01.08.2017 to 31.07.2020 (Three Years)	100 Nos.	---

17/07/17

Date : 17.07.2017



Govt. Registered Engineer

Er. J. PISON JOE, B.Sc. Engg., M.S. Engg., P.E.I.V.

Govt. Registered Engineer

JOE ASSOCIATES

CHIRAYANKUZHI, MARTHANDAM Ph. 04651-27108



JOE ASSOCIATES

CONSULTING ARCHITECTS, ENGINEERS & VALUERS

BERACHAH TOWER, CHIRAYANKUZHI JN., KANJIRACODE P.O, PIN : 629 155.

Ph : (0) 04651 - 271080, (R) 273080 Mob : 9443607580, Email : pisonjoe@yahoo.com.

Er. J. PISON JOE, B.Sc., B.E., M.I.E., C.Eng (I), F.I.V.

Website : www.joeassociates.com

Govt. Regd. Engineer | Govt. Regd. Valuer | Indian Bank Panel Valuer | Karnataka Bank Panel Valuer | Federal Bank Panel Valuer |
Chartered Engineer | Approved Valuer | H.D.F.C. Panel Valuer | Corporation Bank Panel Valuer | S.B.I. Panel Valuer | L.I.C. Panel Valuer |

STATEMENT ACCOMPANYING FORM - A

S30/C/17

1. Door Number of Building : 3 - 101 / 4
2. Location (street no. ward no.) Name of place i.e. Corp./ Municipality/ Panchayat Union / Township : Sarada Krishna Homeopathic Medical College Ladies Hostel, Kulasekharam P.O. R.S. Nos. 392 / 2A, 2B2
3. Year of Construction : Partly 2006 & 2011
4. Approximate cost of construction as per current rates : Rs. 97,60,000/-
5. Purpose for which the building is being used or proposed to be used. : Ladies Hostel
6. Purpose for which the building has hitherto been used : Ladies Hostel
7. Details of construction of the building (particulars of the foundation soil, specification of the various parts under):
 - i. Foundation : R.C.C. Column Footing & R.R.M in C. M.
 - ii. Basement : R.C.C. Column & R.R.M. in C.M.
 - iii. Superstructure : R.C.C. Column & B.W. in C.M.
 - iv. Floor : Marble & Marbonite Flooring
 - v. Roof : R.C.C.
8. Period for which license is required : 01.08.2017 to 31.07.2020 (Three Years)
9. Number of persons to be accommodated : 500 Nos.

Date: 17.07.2017

Govt. Registered Engineer

Er. J. PISON JOE, B.Sc., B.E., M.I.E., C.Eng (I), F.I.V.

Govt. Registered Engineer

JOE ASSOCIATES

CHIRAYANKUZHI, MARTHANDAM Ph. 04651-271080.

Certified that no addition or alteration has been made to the existing building or portion of the building for which this application has been made. I undertake to obtain fresh license, in case addition or alteration is made to the existing building or portion of the building. I also undertake to obtain fresh license, if the purpose for which the license was granted for the use of the building originally is changed or altered subsequently.

Date:

Signature of the owner of the building

FORM - B

Form of certificate of Structural Soundness under sub-section 4 of the Madras Public Buildings (Licensing) Act 1965, (Madras Act 13 of 1965) referred to in Rule of the Madras Public Building Rules 1966.

I certify that I have inspected the buildings mentioned in the statement and furnish below its technical features:

1. FOUNDATION

- (A)
- i. Depth below Ground Level : 1.95 m
 - ii. Nature of soil met with at foundation level : Hard Gravelly Soil
 - iii. Pressure of foundation : 34.43 t/m²
 - iv. Probable S.B Capacity of the strata met with at foundation : 45 t/m²

2. SUPERSTRUCTURE

- (A)
- i. Maximum stress in the most critical section of masonry and Concrete : Does not arise since it is a framed structure
 - ii. Nature of masonry and Concrete and its safe permissible stress : —

The structural soundness of the building has been verified by me with reference to ISI loading Standard 875 (Latest version) and other relevant INDIAN STANDARDS CODE OF PRACTICE and I declare the building is structurally sound to be used as Public Building for the purpose noted in the statement. A license may be granted for the period from **01.08.2017 to 31.07.2020 (Three Years)**

(A)

M/102/17

Date: 17.07.2017


Govt. Registered Engineer

Er. J. PISON JCE, B.Sc. D.E., M.E. (Civ.), F.I.V.
Govt. Registered Engineer
JOE ASSOCIATES
CHIRAYANKUZZHI, MARTHANDAM Ph. 04051-271030.

FORM - C

STATEMENT TO ACCOMPANY THE CERTIFICATE OF STRUCTURAL SOUNDNESS

Location of the Building Door No., Street and Place.	Age of the building	No. of stories	Short description	Certified purpose the building stability	Certified period for the building stability	No. of persons to be accommodated	Remarks
D. No. 3 - 101 / 4 Sarada Krishna Homeopathic Medical College, Ladies Hostel, Kulasekharam P.O. R.S. Nos. 392 / 2A, 2B2	10 Years	G.F., F. F. S. F. & T. F. Four stories	R.C.C. Framed Structure, Country wood doors, windows,	Ladies Hostel	01.08.2017 to 31.07.2020 (Three Years)	500 Nos.	---



Date : 17.07.2017

Govt. Registered Engineer

Er. J. PISON JOE, B.Sc., B.E., M.I.E., C.Eng(I), F.I.V.

Govt. Registered Engineer

JOE ASSOCIATES

CHIRAYANKUZHI, MARTHANDAM Ph. 04651-271080.



JOE ASSOCIATES

CONSULTING ARCHITECTS, ENGINEERS & VALUERS

BERACHAH TOWER, CHIRAYANKUZHI Jn., KANJIRACODE P.O, PIN : 629 155.

Ph : (O) 04651 - 271080, (R) 273080 Mob : 9443607580, Email : pisonjoe@yahoo.com.

Er. J. PISON JOE, B.Sc., B.E., M.I.E., C.Eng(I), F.I.V.

Website : www.joeassociates.com

Govt. Regd. Engineer | Govt. Regd. Valuer | Income Tax Valuer | Indian Bank Panel Valuer | Karnataka Bank Panel Valuer | Federal Bank Panel Valuer | Chartered Engineer | Approved Valuer | H.D.F.C. Panel Valuer | Corp. Bank Panel Valuer | S.B.I. Panel Valuer | S.B.T. Panel Valuer | L.I.C. Panel Valuer

22 - 03 - 2017

Completion Certificate

This is to certify that, the construction of the Hostel building bearing Door No. 3 – 101 / 4, which is used for Sarada Krishna Homeopathic Medical College Ladies Hostel, in R.S. Nos.: 392 / 2 A & 2 B 2 at Kulasekharam, is fully completed during the year 2011, in all aspects and fit for use and occupation.

22/03/17

Signature

Er. J. PISON JOE, B.Sc., B.E., M.I.E., C.Eng(I), F.I.V.
Govt. Registered Engineer
JOE ASSOCIATES
CHIRAYANKUZHI, MARTHANDAM Ph. 04651-271080.



JOE ASSOCIATES

CONSULTING ARCHITECTS, ENGINEERS & VALUERS
BERACHAH TOWER, CHIRAYANKUZHI Jn., KANJIRACODE P.O, PIN : 629 155.

Ph : (O) 04651 - 271080, (R) 273080 Mob : 9443607580, Email : pisonjoe@yahoo.com.

Er. J. PISON JOE, B.Sc., B.E., M.I.E., C.Eng(I), F.I.V.

Website : www.joeassociates.com

Govt. Regd. Engineer | Govt. Regd. Valuer | Income Tax Valuer | Indian Bank Panel Valuer | Karnataka Bank Panel Valuer | Federal Bank Panel Valuer |
Chartered Engineer | Approved Valuer | H.D.F.C. Panel Valuer | Corp. Bank Panel Valuer | S.B.I. Panel Valuer | S.B.T. Panel Valuer | L.I.C. Panel Valuer

22 - 03 - 2017

Completion Certificate

This is to certify that, the construction of the Hostel building bearing Door No. 3 – 101 / 1, which is used for **Sarada Krishna Homeopathic Medical College Gents Hostel**, in R.S. No.: 393 / 7 at Kulasekharam, is **fully completed** during the year 2008, in all aspects and fit for use and occupation.




Signature

Er. J. PISON JOE, B.Sc., B.E., M.I.E., C.Eng(I), F.I.V.
Govt. Registered Engineer
JOE ASSOCIATES
CHIRAYANKUZHI, MARTHANDAM PH. 04651-271080.



JOE ASSOCIATES

CONSULTING ARCHITECTS, ENGINEERS & VALUERS

BERACHAH TOWER, CHIRAYANKUZZHI Jn., KANJIRACODE P.O, PIN : 629 155.

Ph : (O) 04651 - 271080, (R) 273080 Mob : 9443607580, Email : pisonjoe@yahoo.com.

Er. J. PISON JOE, B.Sc., B.E., M.I.E., C.Eng(I), F.I.V.

Website : www.joeassociates.com

Govt. Regd. Engineer | Govt. Regd. Valuer | Income Tax Valuer | Indian Bank Panel Valuer | Karnataka Bank Panel Valuer | Federal Bank Panel Valuer |
Chartered Engineer | Approved Valuer | H.D.F.C. Panel Valuer | Corp. Bank Panel Valuer | S.B.I. Panel Valuer | S.B.T. Panel Valuer | L.I.C. Panel Valuer

22 - 03 - 2017

Completion Certificate

This is to certify that, the construction of the College Hospital building bearing Door Nos. 3 – 101 / 5 & 6, which is used for Sarada Krishna Homeopathic Medical College Hospital, in R.S. Nos.: 392 / 2 A & 2 B at Kulasekharam, is fully completed during the year 2006, in all aspects and fit for use and occupation.

22/03/17


Signature

Er. J. PISON JOE, B.Sc., B.E., M.I.E., C.Eng(I), F.I.V.
Govt. Registered Engineer
JOE ASSOCIATES
CHIRAYANKUZZHI, MARTHANDAM Ph. 04651-271080



JOE ASSOCIATES

CONSULTING ARCHITECTS, ENGINEERS & VALUERS

BERACHAH TOWER, CHIRAYANKUZHI Jn., KANJIRACODE P.O, PIN : 629 155.

Ph : (O) 04651 - 271080, (R) 273080 Mob : 9443607580, Email : pisonjoe@yahoo.com.

Er. J. PISON JOE, B.Sc., B.E., M.I.E., C.Eng(I), F.I.V.

Website : www.joeassociates.com

Govt. Regd. Engineer | Govt. Regd. Valuer | Income Tax Valuer | Indian Bank Panel Valuer | Karnataka Bank Panel Valuer | Federal Bank Panel Valuer |
Chartered Engineer | Approved Valuer | H.D.F.C. Panel Valuer | Corp. Bank Panel Valuer | S.B.I. Panel Valuer | S.B.T. Panel Valuer | L.I.C. Panel Valuer

22 - 03 - 2017

Completion Certificate

This is to certify that, the construction of the college building bearing Door Nos. 3 - 101, 3 - 101 / 2 & 3, which is used for Sarada Krishna Homeopathic Medical College, in R.S. Nos.: 392 / 2 A & 2 B 2 at Kulasekharam, is fully completed during the year 2000, in all aspects and fit for use and occupation




Signature

Er. J. PISON JOE, B.Sc., B.E., M.I.E., C.Eng(I), F.I.V.
Govt. Registered Engineer
JOE ASSOCIATES
CHIRAYANKUZHI, MARTHANDAM Ph. 04651-271080.

From
Thiru V. Pandian, I.A.S.,
District Collector,
Kanniyakumari District,
Nagarcoll.

To
Mr. J. Pison Joo, B.Sc., B.E., H.J.E.,
C.Engg.
Rotha Paper Store Building,
Marthanda 629165

E1/88538/97 dt. 11-11-98

Sir,


Re: Inclusion of the name of private practising
Engineer Class-I, confirmation letter requested
regarding.

Re: Your letter dt. 30-10-98

It is hereby informed that your name has been included
in the panel of Private Practising Engineers under class-I in
Kanniyakumari District as per letter No ABB/71/JE/41931/98 dated
6.7.98 of the Engineer-in-Chief, WAO & C.E. (General), P.W.D.
Chennai.

/t.c. s.o./

Yours Faithfully,
A. Natarajan
for Collector


Head Assistant (B) 1314/1

TAMIL NADU FIRE – RESCUE SERVICE

FIRE –LICENSE

(See Section 13)

L.Dis.No:2413/ A / 2019

Date : 22 .04.2019

Licence is hereby granted under section 13 of the Tamil Nadu Fire Service Act. 1985 to for Running Homoeopathic Medical College in the name of "M/s.Sarada Krishna Homoeopathic Medical College" within the Jurisdiction kalkulam Taluk at the Premises Door No.3-101, 3-101/ 2 & 3, 3-101/5 & 6 in Survey No.392/2A, 2B2 of Kulasekaram "B" Village, Kalkulam Taluk, Kanyakumari District. Subject to the Condition noted thereon and such other condition as may be prescribed.

This Fire Licence is valid for one year from the date of issue of this license.

CONDITIONS:

- 1) *The Installed Fire Fighting Equipments should be maintained in good condition at all times.*
- 2) *All exits should be free of obstructions. Exits should be clearly visible and the routes to reach the exit shall be clearly marked.*
- 3) *Trained Fire personnel should be available to operate the Fire Fighting Equipments in case of any emergency*
- 4) *"Mock Drill Should be conducted periodically and the date of conduction should be informed to this office well in advance to enable the team to make suggestions if any*
- 5) *Dumping of waste materials anywhere should be avoided. Dust bin should be cleared frequently*
- 6) *Any erection of permanent or temporary structure should be intimated to this department.*
- 7) *Do s Don'ts in time of emergencies should be available in a laminated hanging card.*
- 8) *Permission should be obtained from the Local Planning Authority/ Town and Country Planning Authority*

Office Seal with Date



District Officer
Fire & Rescue Services
Kanyakumari District, Nagercoil

To The Principal,
M/s.Sarada Krishna Homoeopathic Medical College
Door No.3-101, 3-101/ 2 & 3, 3-101/5 & 6
Kulasekaram Village,
Kalkulam Taluk,
Kanyakumari District. 628 161

Handwritten signature and date 22/4/19.

Proceedings of the Tahsildar, Kalkulam
Present: Thiru.C.Raja., B.A.,



K.Dis. A6/5188/2018

Dated : 18.05.2018

3392

Sub: Building Licence – Kalkulam Taluk – Kulasekharam 'B'
 Village – Saradha Krishna Homeopathic Medical College &
 Hospital, Kulasekharam - renewal of building Licence - order
 issued - regarding.
 Ref: Connected records.

Order:

Building Licence is favour of the Principal, Saradha Krishna Homeopathic Medical College & Hospital, Kulasekharam for the College Building bearing Door No. 3/101, 3-101/2&3 and the Hospital Building bearing Door No. 3-101/5&6 of Kulasekharam Town Panchayat. The Licence is valid for the period from 01.04.2018 to 31.03.2021. The Licence is issued under section 2 (i) of section 7 of the Tamil Nadu Building Licence Act, 1965, subject of the following conditions.

1. Saradha Krishna Homeopathic Medical College & Hospital, Kulasekharam authorities should not construct any roofs with materials which are easily inflammable and pose threat of fire menace after receipt of building licence.
2. They should maintain electrical connections properly and it should be fire free.
3. They should pave way to impart trainings to their staffs and enable them to do firefighting demonstrations in the event of contingencies.
4. They are solely responsible for any lapse on above lines if any fire mishap takes place in future.

The licence in form D is enclosed herewith.

Encl:- Form D.

Tahsildar,
 Kalkulam.

To
 The Principal,
 Saradha Krishna Homeopathic Medical College & Hospital,
 Kulasekharam.

21/5/18

FORM - D

Form of Licence under section (1) of Section 6 of the Madras Public Building (Licencing) ACT 1965 (Madras Act 19 of 1965) referred to the rule of the Madras Public Building Licencing Rules 1965.

Registration No. A6/5188/2018

License No. 34/2018
Fee Rs. 5000/-

Building Licence is here by granted Saradha Krishna Homeopathic Medical College & Hospital, Kulasekharam for the purpose and in respect of the building in the statement below and sent to the conditions and for the period mentioned here under.

Door.No.

Period

The period of validity of the Licence shall be from

3/101, 3-101/2&3
3-101/5&6

01.04.2018 to 31.03.2021
01.04.2018 to 31.03.2021



Tahsildar,
Kalkulam.

STATEMENT TO ACCOMPANY THE LICENSE

Location of the building (Door No. Street No. and Name of Place)	Purpose for which the building is licenced to be used as a public buildings	No. of persons to be accommodated
D.No. 3/101, 3-101/2&3		700
D.No. 3-101/5&6 of Kulasekharam Town Panchayat.	Educational Hospital	500

Tahsildar,
Kalkulam



TransCal

Measurement to Perfection...

Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003



NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Sarada Krishna Homoeopathic Medical College Hospital
Kulashekaram, Kanyakumari District, Tamil Nadu-629161

Customer's Reference :

SRF No. : TSC/18-19/7168

Dated : 12 Oct 2018

ULR.NO CC223118000004186F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/7168-10	13 Oct 2018	13 Oct 2019	1 of 2

Details of device under calibration

Transcal ID : TSC142342

Nomenclature : Height Scale	No. of Pages : 2
Make : Biocon	Cal Procedure No. : TSC/CAL/428
Model/Range : 0-200cm	DUC Received : 12 Oct 2018
SI No. : NA	DUC Condition on Receipt : Satisfactory
ID No. : SKHMC/BM/H S-003	Cal At : Dimensional Lab

Environmental Conditions : Temperature in °C : 20.2

Humidity in RH % : 52.1

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Length Measuring Machine	ME	0-1000 mm	010300	CGIL/2017/0422/C-001	21 Apr 2019

Note :

- This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
- Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
- Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
- Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
- Calibration of the DUC are traceable to National/International Standards
- Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data (To be filled by customer authorized signatory and not under calibration laboratory control).
- In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
- Consider Model or Range whichever is applicable.
- Nabl-133 guidelines are adopted for use of NABL symbol.



Calibrated By

Anandesh T N
(Calibration Engineer)

Checked By

Pavan P K
(Calibration Engineer)

Authorised By

Anandmurthy B.S
(Lab Incharge)



CAL CERT. NO : TSC/18-19/7168-10

Page : 2 of 2

Range: 0 - 2000 mm

L.C.: 1 mm

Results:

Sl. No.	DUC Marking in mm	STD Readings in mm	Error Observed in mm
1	150	150.000	0.000
2	200	199.975	0.025
3	300	300.075	-0.075
4	500	500.780	-0.780
5	750	751.410	-1.410
6	1000	1002.160	-2.160
7	1500	1503.135	-3.135
8	2000	2004.240	-4.240

Conclusion/ Remarks:

Limits of Tolerance as per IS:1269-1997

Class	Maximum permissible error for initial verification
Class I	$\pm (0.1 + 0.1L)$ mm
Class II	$\pm (0.3 + 0.2L)$ mm
Class III	$\pm (0.6 + 0.4L)$ mm

"L" is length of the measuring tape in meter

Maximum permissible error for Measures in Service is Equal to twice the maximum permissible error at the the time of initial verification.

Measurement Uncertainty is $\pm (565 + L/100)$ μ m at 95% Confidence level , with $k=1.96$



Calibrated By

Anandesh T N
(Calibration Engineer)

Checked By

Pavan P K
(Calibration Engineer)

Authorised By

Anandmurthy B.S
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Sarada Krishna Homoeopathic Medical College Hospital
Kulashhekaram, Kanyakumari District, Tamil Nadu-629161

Customer's Reference : SRF No. : TSC/18-19/7168 Dated : 12 Oct 2018

ULR.NO CC223118000004179F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/7168-11	13 Oct 2018	13 Oct 2019	1 of 2

Details of device under calibration Transcal ID : TSC142343

Nomenclature : Height Scale	No. of Pages : 2
Make : Biocon	Cal Procedure No. : TSC/CAL/428
Model/Range : 0-200cm	DUC Received : 12 Oct 2018
SI No. : NA	DUC Condition on Receipt : Satisfactory
ID No. : SKHMC/BM/H S-007	Cal At : Dimensional Lab

Environmental Conditions : Temperature in °C : 20.1

Humidity in RH % : 52.1

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Length Measuring Machine	ME	0-1000 mm	010300	CGIL/2017/0422/C-001	21 Apr 2019

Note :

- This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
- Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
- Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
- Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
- Calibration of the DUC are traceable to National/International Standards
- Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
- In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
- Consider Model or Range whichever is applicable.
- Nabl-133 guidelines are adopted for use of NABL symbol.



Calibrated By

Anandesh T N
(Calibration Engineer)

Checked By

Pavan P K
(Calibration Engineer)

Authorised By

Anandmurthy B.S
(Lab Incharge)



CAL CERT. NO : TSC/18-19/7168-11

Page : 2 of 2

Range: 0 - 2000 mm

L.C.: 1 mm

Results:

Sl. No.	DUC Marking in mm	STD Readings in mm	Error Observed in mm
1	150	150.000	0.000
2	200	199.980	0.020
3	300	300.050	-0.050
4	500	500.715	-0.715
5	750	751.095	-1.095
6	1000	1001.360	-1.360
7	1500	1502.065	-2.065
8	2000	2002.750	-2.750

Conclusion/ Remarks:

Limits of Tolerance as per IS:1269-1997

Class	Maximum permissible error for intial verification
Class I	$\pm (0.1 + 0.1L)$ mm
Class II	$\pm (0.3 + 0.2L)$ mm
Class III	$\pm (0.6 + 0.4L)$ mm

"L" is length of the measuring tape in meter

Maximum permissible error for Measures in Service is Equal to twice the maximum permissible error at the the time of intial verification.

Measurement Uncertainty is $\pm (565 + L/100)$ μ m at 95% Confidence level , with k=1.96




Calibrated By


Anandesh T N
(Calibration Engineer)

Checked By


Pavan P K
(Calibration Engineer)

Authorised By


Anandmurthy B.S
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Sarada Krishna Homoeopathic Medical College Hospital
Kulashekaram, Kanyakumari District, Tamil Nadu-629161

Customer's Reference : SRF No. : TSC/18-19/7168 Dated : 12 Oct 2018
ULR.NO CC223118000004180F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/7168-12	13 Oct 2018	13 Oct 2019	1 of 2

Details of device under calibration		Transcal ID	
Nomenclature	: Height Scale	No. of Pages	: 2
Make	: Biocon	Cal Procedure No.	: TSC/CAL/428
Model/Range	: 0-200cm	DUC Received	: 12 Oct 2018
SI No.	: NA	DUC Condition on Receipt	: Satisfactory
ID No.	: SKHMC/BM/H S-008	Cal At	: Dimensional Lab

Environmental Conditions : Temperature in °C : 20.2

Humidity in RH % : 52.0

Standards used :

SI No.	Nomenclature	Make	Model	SI No/ID.No.	Certificate No.	Validity
1	Length Measuring Machine	ME	0-1000 mm	010300	CGIL/2017/0422/C-001	21 Apr 2019

Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data (To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By

Anandesh T N
(Calibration Engineer)

Checked By

Pavan P K
(Calibration Engineer)



Authorised By

Anandmurthy B.S
(Lab Incharge)



CAL CERT. NO : TSC/18-19/7168-12

Page : 2 of 2

Range: 0 - 2000 mm

L.C.: 1 mm

Results:

Sl. No.	DUC Marking in mm	STD Readings in mm	Error Observed in mm
1	150	150.000	0.000
2	200	200.020	-0.020
3	300	300.270	-0.270
4	500	501.105	-1.105
5	750	751.990	-1.990
6	1000	1002.680	-2.680
7	1500	1504.510	-4.510
8	2000	2005.555	-5.555

Conclusion/ Remarks:

Limits of Tolerance as per IS:1269-1997

Class	Maximum permissible error for intial verification
Class I	$\pm (0.1 + 0.1L)$ mm
Class II	$\pm (0.3 + 0.2L)$ mm
Class III	$\pm (0.6 + 0.4L)$ mm

"L" is length of the measuring tape in meter


Maximum permissible error for Measures in Service is Equal to twice the maximum permissible error at the the time of intial verification.

Measurement Uncertainty is $\pm (565 + L/100)$ μ m at 95% Confidence level , with k=1.96

Calibrated By



Anandesh T N
(Calibration Engineer)

Checked By


Pavan P K
(Calibration Engineer)



Authorised By


Anandmurthy B.S
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At :100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Sarada Krishna Homoeopathic Medical College Hospital
Kulashkaram, Kanyakumari District, Tamil Nadu-629161

Customer's Reference :

SRF No. : TSC/18-19/7168

Dated : 12 Oct 2018

ULR.NO CC223118000004185F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/7168-13	13 Oct 2018	13 Oct 2019	1 of 2

Details of device under calibration

Transcal ID : TSC142346

Nomenclature : Height Scale	No. of Pages : 2
Make : Biocon	Cal Procedure No. : TSC/CAL/428
Model/Range : 0-200cm	DUC Received : 12 Oct 2018
SI No. : NA	DUC Condition on Receipt : Satisfactory
ID No. : --	Cal At : Dimensional Lab

Environmental Conditions : Temperature in °C : 19.9

Humidity in RH % : 52.0

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Length Measuring Machine	ME	0-1000 mm	010300	CGIL/2017/0422/C-001	21 Apr 2019


Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data (To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By



Anandesh T N
(Calibration Engineer)

Checked By


Pavan P K
(Calibration Engineer)



Authorised By


Anandmurthy B.S
(Lab Incharge)



CAL CERT. NO : TSC/18-19/7168-13

Page : 2 of 2

Range: 0 - 2000 mm

L.C.: 1 mm

Results:

Sl. No.	DUC Marking in mm	STD Readings in mm	Error Observed in mm
1	150	150.000	0.000
2	200	200.125	-0.125
3	300	299.865	0.135
4	500	499.625	0.375
5	750	749.830	0.170
6	1000	1000.190	-0.190
7	1500	1500.120	-0.120
8	2000	1999.915	0.085

Conclusion/ Remarks:

Limits of Tolerance as per IS:1269-1997

Class	Maximum permissible error for initial verification
Class I	$\pm (0.1 + 0.1L)$ mm
Class II	$\pm (0.3 + 0.2L)$ mm
Class III	$\pm (0.6 + 0.4L)$ mm

"L" is length of the measuring tape in meter


Maximum permissible error for Measures in Service is Equal to twice the maximum permissible error at the the time of initial verification.

Measurement Uncertainty is $\pm (565 + L/100)$ μ m at 95% Confidence level , with k=1.96

Calibrated By



Anandesh T N
(Calibration Engineer)

Checked By


Pavan P K
(Calibration Engineer)



Authorised By


Anandmurthy B.S
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Sarada Krishna Homoeopathic Medical College Hospital
Kulashekaram, Kanyakumari District, Tamil Nadu-629161

Customer's Reference : SRF No. : TSC/18-19/7168 Dated : 12 Oct 2018
ULR.NO CC223118000004182F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/7168-14	13 Oct 2018	13 Oct 2019	1 of 2

Details of device under calibration		Transcal ID	
Nomenclature	: Height Scale	No. of Pages	: 2
Make	: Biocon	Cal Procedure No.	: TSC/CAL/428
Model/Range	: 0-200cm	DUC Received	: 12 Oct 2018
SI No.	: NA	DUC Condition on Receipt	: Satisfactory
ID No.	: SKHMC/BM/H S 005	Cal At	: Dimensional Lab

Environmental Conditions : Temperature in °C : 19.9 Humidity in RH % : 52.

Standards used :

SI No.	Nomenclature	Make	Model	SI No/ID.No.	Certificate No.	Validity
1	Length Measuring Machine	ME	0-1000 mm	010300	CGIL/2017/0422/C-001	21 Apr 2019

Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.



Calibrated By

Anandesh T N
(Calibration Engineer)

Checked By

Pavan P K
(Calibration Engineer)

Authorised By

Anandmurthy B.S
(Lab Incharge)



CAL CERT. NO : TSC/18-19/7168-14

Page : 2 of 2

Range: 0 - 2000 mm

L.C.: 1 mm

Results:

Sl. No.	DUC Marking in mm	STD Readings in mm	Error Observed in mm
1	150	150.000	0.000
2	200	200.005	-0.005
3	300	300.010	-0.010
4	500	500.070	-0.070
5	750	750.035	-0.035
6	1000	1000.020	-0.020
7	1500	1499.505	0.495
8	2000	1999.515	0.485

Conclusion/ Remarks:

Limits of Tolerance as per IS:1269-1997

Class	Maximum permissible error for intial verification
Class I	$\pm (0.1 + 0.1L)$ mm
Class II	$\pm (0.3 + 0.2L)$ mm
Class III	$\pm (0.6 + 0.4L)$ mm

"L" is length of the measuring tape in meter


Maximum permissible error for Measures in Service is Equal to twice the maximum permissible error at the the time of intial verification.

Measurement Uncertainty is $\pm (565 + L/100)$ μ m at 95% Confidence level , with k=1.96

Calibrated By



Anandesh T N
(Calibration Engineer)

Checked By


Pavan P K
(Calibration Engineer)



Authorised By


Anandmurthy B.S
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Sarada Krishna Homoeopathic Medical College Hospital
Kulashekaram, Kanyakumari District, Tamil Nadu-629161

Customer's Reference : SRF No. : TSC/18-19/7168 Dated : 12 Oct 2018
ULR.NO CC223118000004184F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/7168-15	13 Oct 2018	13 Oct 2019	1 of 2

Details of device under calibration		Transcal ID : TSC142356	
Nomenclature	: Height Scale	No. of Pages	: 2
Make	: Biocon	Cal Procedure No.	: TSC/CAL/428
Model/Range	: 0-200cm	DUC Received	: 12 Oct 2018
SI No.	: NA	DUC Condition on Receipt	: Satisfactory
ID No.	: SKHMC/BM/H S 009	Cal At	: Dimensional Lab

Environmental Conditions : Temperature in °C : 20.1 Humidity in RH % : 51.8

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Length Measuring Machine	ME	0-1000 mm	010300	CGIL/2017/0422/C-001	21 Apr 2019

Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.



Calibrated By

Anandesh T N
(Calibration Engineer)

Checked By

Pavan P K
(Calibration Engineer)

Authorised By

Anandmurthy B.S
(Lab Incharge)



CAL CERT. NO : TSC/18-19/7168-15

Page : 2 of 2

Range: 0 - 2000 mm

L.C.: 1 mm

Results:

Sl. No.	DUC Marking in mm	STD Readings in mm	Error Observed in mm
1	150	150.000	0.000
2	200	199.935	0.065
3	300	300.190	-0.190
4	500	500.875	-0.875
5	750	751.385	-1.385
6	1000	1002.080	-2.080
7	1500	1503.295	-3.295
8	2000	2004.285	-4.285

Conclusion/ Remarks:

Limits of Tolerance as per IS:1269-1997

Class	Maximum permissible error for initial verification
Class I	$\pm (0.1 + 0.1L)$ mm
Class II	$\pm (0.3 + 0.2L)$ mm
Class III	$\pm (0.6 + 0.4L)$ mm

"L" is length of the measuring tape in meter


Maximum permissible error for Measures in Service is Equal to twice the maximum permissible error at the the time of initial verification.

Measurement Uncertainty is $\pm (565 + L/100)$ μ m at 95% Confidence level , with k=1.96

Calibrated By



Anandesh T N
(Calibration Engineer)

Checked By


Pavan P K
(Calibration Engineer)



Authorised By


Anandmurthy B.S
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld .Premises : Centenary Building (G. Flr), Door No. At :100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Sarada Krishna Homoeopathic Medical College Hospital
Kulashekaram, Kanyakumari District, Tamil Nadu-629161

Customer's Reference :

SRF No. : TSC/18-19/7168

Dated : 12 Oct 2018

ULR.NO CC223118000004192F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/7168-16	13 Oct 2018	13 Oct 2019	1 of 2

Details of device under calibration

Transcal ID : TSC142359

Nomenclature : Height Scale	No. of Pages : 2
Make : Biocon	Cal Procedure No. : TSC/CAL/428
Model/Range : 0-200cm	DUC Received : 12 Oct 2018
SI No. : NA	DUC Condition on Receipt : Satisfactory
ID No. : SKHMC/BM/H S 006	Cal At : Dimensional Lab

Environmental Conditions : Temperature in °C : 19.8

Humidity in RH % : 52.1

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Length Measuring Machine	ME	0-1000 mm	010300	CGIL/2017/0422/C-001	21 Apr 2019

Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.



Calibrated By

Anandesh T N
(Calibration Engineer)

Checked By

Pavan P K
(Calibration Engineer)

Authorised By

Anandmurthy B.S
(Lab Incharge)



CAL CERT. NO : TSC/18-19/7168-16

Page : 2 of 2

Range: 0 - 2000 mm

L.C.: 1 mm

Results:

Sl. No.	DUC Marking in mm	STD Readings in mm	Error Observed in mm
1	150	150.000	0.000
2	200	199.920	0.080
3	300	300.220	-0.220
4	500	500.890	-0.890
5	750	751.195	-1.195
6	1000	1001.620	-1.620
7	1500	1502.390	-2.390
8	2000	2003.375	-3.375

Conclusion/ Remarks:

Limits of Tolerance as per IS:1269-1997

Class	Maximum permissible error for intial verification
Class I	$\pm (0.1 + 0.1L)$ mm
Class II	$\pm (0.3 + 0.2L)$ mm
Class III	$\pm (0.6 + 0.4L)$ mm

"L" is length of the measuring tape in meter

Maximum permissible error for Measures in Service is Equal to twice the maximum permissible error at the the time of intial verification.

Measurement Uncertainty is $\pm (565 + L/100)$ μ m at 95% Confidence level , with k=1.96



Calibrated By


Anandesh T N
(Calibration Engineer)

Checked By


Pavan P K
(Calibration Engineer)

Authorised By


Anandmurthy B.S
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld.Premises : Centenary Building (G.Fir), Door No. At :100, W.Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Sarada Krishna Homoeopathic Medical College Hospital
Kulashekaram, Kanyakumari District, Tamil Nadu-629161

Customer's Reference :

SRF No. : TSC/18-19/7168

Dated : 12 Oct 2018

ULR.NO CC223118000004177F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/7168-17	13 Oct 2018	13 Oct 2019	1 of 2

Details of device under calibration

Transcal ID : TSC142346

Nomenclature : Height Scale	No. of Pages : 2
Make : Biocon	Cal Procedure No. : TSC/CAL/428
Model/Range : 0-200cm	DUC Received : 12 Oct 2018
SI No. : NA	DUC Condition on Receipt : Satisfactory
ID No. : --	Cal At : Dimensional Lab

Environmental Conditions : Temperature in °C : 19.8

Humidity in RH % : 52.0

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Length Measuring Machine	ME	0-1000 mm	010300	CGIL/2017/0422/C-001	21 Apr 2019

Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.



Calibrated By

Anandesh T N
(Calibration Engineer)

Checked By

Pavan P K
(Calibration Engineer)

Authorised By

Anandmurthy B.S
(Lab Incharge)



CAL CERT. NO : TSC/18-19/7168-17

Page : 2 of 2

Range: 0 - 2000 mm

L.C.: 1 mm

Results:

Sl. No.	DUC Marking in mm	STD Readings in mm	Error Observed in mm
1	150	150.000	0.000
2	200	199.985	0.015
3	300	299.895	0.105
4	500	499.565	0.435
5	750	749.815	0.185
6	1000	999.315	0.685
7	1500	1499.225	0.775
8	2000	1999.154	0.846

Conclusion/ Remarks:

Limits of Tolerance as per IS:1269-1997

Class	Maximum permissible error for intial verification
Class I	$\pm (0.1 + 0.1L)$ mm
Class II	$\pm (0.3 + 0.2L)$ mm
Class III	$\pm (0.6 + 0.4L)$ mm

"L" is length of the measuring tape in meter


Maximum permissible error for Measures in Service is Equal to twice the maximum permissible error at the the time of intial verification.

Measurement Uncertainty is $\pm (565 + L/100)$ μ m at 95% Confidence level , with $k=1.96$

Calibrated By



Anandesh T N
(Calibration Engineer)

Checked By


Pavan P K
(Calibration Engineer)



Authorised By


Anandmurthy B.S
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises: Centenary Building (G. Flr), Door No. At :100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Sarada Krishna Homoeopathic Medical College Hospital
Kulashekaram, Kanyakumari District, Tamil Nadu-629161

Customer's Reference :

SRF No. : TSC/18-19/7168

Dated : 12 Oct 2018

ULR.NO CC22311800004190F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/7168-7	22 Oct 2018	22 Apr 2019	1 of 2

Details of device under calibration

Transcal ID : TSC82457

Nomenclature : Micro Pipette	No. of Pages : 2
Make : Thermo Scientific	Cal Procedure No. : TSC/CAL/610
Model/Range : 100-1000µl	DUC Received : 12 Oct 2018
SI No. : NW 08991	DUC Condition on Receipt : Satisfactory
ID No. : --	Cal At : Mechanical Lab

Environmental Conditions : Temperature in °C : 20.8

Humidity in RH % : 49

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Electronic Weighing Balance	Mettler Toledo	AG 285	1120102251	TSC/17-18/INH/MECH-19	10 Feb 2019

Note :


1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By

Janardhan S
(Calibration Engineer)

Checked By

Chitragadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas B V
(Lab Incharge)



Range : 100-1000 μ l
 Increment : 1 μ l

ATM Pressure : 909.7 hpa

Sl. No.	Micropipette Set Volume in μ l	Standard Balance Reading in g	Actual Calculated Volume @ 27°C in μ l	Average Volume in μ l	Systematic Error, \pm in %	Random Error, in \pm in %
1	100	0.10021	100.51	100.65	0.65	0.13
2		0.10025	100.55			
3		0.10026	100.56			
4		0.10030	100.60			
5		0.10034	100.64			
6		0.10039	100.69			
7		0.10045	100.75			
8		0.10052	100.82			
9		0.10058	100.88			
10		0.10019	100.49			
11	500	0.50478	506.29	506.50	1.30	0.03
12		0.50483	506.34			
13		0.50489	506.40			
14		0.50496	506.47			
15		0.50503	506.54			
16		0.50507	506.58			
17		0.50512	506.63			
18		0.50519	506.70			
19		0.50524	506.75			
20		0.50476	506.27			
21	1000	1.00631	1009.33	1009.50	0.95	0.01
22		1.00638	1009.40			
23		1.00642	1009.44			
24		1.00647	1009.49			
25		1.00652	1009.54			
26		1.00655	1009.57			
27		1.00657	1009.59			
28		1.00664	1009.66			
29		1.00670	1009.72			
30		1.00629	1009.31			

Measurement Uncertainty : \pm 0.20 μ l

Conclusion / Remarks:

- 1 Measurement uncertainty is at confidence level 95% which corresponds to a coverage factor of k= 2.26
- 2 Calibration is performed as per ISO 8655 - 6 : 2002 (E)
- 3 Gravimetric Method is adopted for calibration

Calibrated By

[Signature]
Janardhan S
 (Calibration Engineer)

Checked By

[Signature]
Chitrangadha P R
 (Sr. Calibration Engineer)



Authorised By

[Signature]
Shreyas B V
 (Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Sarada Krishna Homoeopathic Medical College Hospital
Kulashekaram, Kanyakumari District, Tamil Nadu-629161

Customer's Reference :

SRF No. : TSC/18-19/7168

Dated : 12 Oct 2018

ULR.NO CC223118000004189F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/7168-6	22 Oct 2018	22 Apr 2019	1 of 2

Details of device under calibration

Transcal ID : TSC46517

Nomenclature : Micro Pipette	No. of Pages : 2
Make : Astra	Cal Procedure No. : TSC/CAL/610
Model/Range : 100-1000µl	DUC Received : 12 Oct 2018
SI No. : DY03443	DUC Condition on Receipt : Satisfactory
ID No. : SKHMC/BM/PPT-009	Cal At : Mechanical Lab

Environmental Conditions : Temperature in °C : 20.7

Humidity in RH % : 49

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Electronic Weighing Balance	Mettler Toledo	AG 285	1120102251	TSC/17-18/INH/MECH-19	10 Feb 2019
2	Micro Balance	Sartorius	MC 21 S	16406121	TSC/17-18/INH/Mech-24	10 Mar 2019

Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By

Janardhan S
(Calibration Engineer)

Checked By

Chitrangadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas B V
(Lab Incharge)



Range : 100-1000 μ l
 Increment : 5 μ l

ATM Pressure : 909.8 hpa

Sl. No.	Micropipette Set Volume in μ l	Standard Balance Reading in g	Actual Calculated Volume @ 27°C in μ l	Average Volume in μ l	Systematic Error, \pm in %	Random Error, in \pm in %
1	100	0.10007	100.37	100.50	0.50	0.10
2		0.10011	100.41			
3		0.10016	100.46			
4		0.10020	100.50			
5		0.10021	100.51			
6		0.10027	100.57			
7		0.10028	100.58			
8		0.10032	100.62			
9		0.10035	100.65			
10		0.10005	100.35			
11	500	0.50028	501.78	501.90	0.38	0.02
12		0.50031	501.81			
13		0.50036	501.86			
14		0.50037	501.87			
15		0.50040	501.90			
16		0.50044	501.94			
17		0.50049	501.99			
18		0.50053	502.03			
19		0.50058	502.08			
20		0.50026	501.76			
21	1000	1.00028	1003.28	1003.40	0.34	0.01
22		1.00031	1003.31			
23		1.00035	1003.35			
24		1.00037	1003.37			
25		1.00042	1003.42			
26		1.00045	1003.45			
27		1.00049	1003.49			
28		1.00052	1003.52			
29		1.00060	1003.60			
30		1.00026	1003.26			

Measurement Uncertainty : \pm 0.20 μ l


Conclusion / Remarks:

- 1 Measurement uncertainty is at confidence level 95% which corresponds to a coverage factor of k= 2.26
- 2 Calibration is performed as per ISO 8655 - 6 : 2002 (E)
- 3 Gravimetric Method is adopted for calibration

Calibrated By



Janardhan S
 (Calibration Engineer)

Checked By


Chitragadha P R
 (Sr. Calibration Engineer)



Authorised By


Shreyas B V
 (Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Sarada Krishna Homoeopathic Medical College Hospital
Kulashekaram, Kanyakumari District, Tamil Nadu-629161

Customer's Reference :

SRF No. : TSC/18-19/7168

Dated : 12 Oct 2018

ULR.NO CC223118000004178F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/7168-5	22 Oct 2018	22 Apr 2019	1 of 2

Details of device under calibration

Transcal ID : TSC108503

Nomenclature : Micro Pipette	No. of Pages : 2
Make : Thermo Scientific	Cal Procedure No. : TSC/CAL/610
Model/Range : 5-50µl	DUC Received : 12 Oct 2018
SI No. : NW05491	DUC Condition on Receipt : Satisfactory
ID No. : --	Cal At : Mechanical Lab

Environmental Conditions : Temperature in °C : 20.8

Humidity in RH % : 49

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Electronic Weighing Balance	Mettler Toledo	AG 285	1120102251	TSC/17-18/INH/MECH-19	10 Feb 2019
2	Micro Balance	Sartorius	MC 21 S	16406121	TSC/17-18/INH/Mech-24	10 Mar 2019

Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By

Janardhan S
(Calibration Engineer)

Checked By

Chitrangadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas B V
(Lab Incharge)



Range : 5 - 50 μ l
 Increment : 0.1 μ l

ATM Pressure : 909.7 hpa

Sl. No.	Micropipette Set Volume in μ l	Standard Balance Reading in g	Actual Calculated Volume @ 27°C in μ l	Average Volume in μ l	Systematic Error, \pm in %	Random Error, in \pm in %
1	5	0.004689	4.703	4.722	-5.56	0.30
2		0.004697	4.711			
3		0.004699	4.713			
4		0.004705	4.719			
5		0.004711	4.725			
6		0.004718	4.732			
7		0.004719	4.733			
8		0.004723	4.737			
9		0.004727	4.741			
10		0.004689	4.703			
11	25	0.02533	25.41	25.55	2.20	0.43
12		0.02537	25.45			
13		0.02542	25.50			
14		0.02549	25.57			
15		0.02550	25.58			
16		0.02551	25.59			
17		0.02556	25.64			
18		0.02559	25.67			
19		0.02564	25.72			
20		0.02533	25.41			
21	50	0.05046	50.61	50.71	1.43	0.17
22		0.05049	50.64			
23		0.05050	50.65			
24		0.05053	50.68			
25		0.05058	50.73			
26		0.05062	50.77			
27		0.05062	50.77			
28		0.05067	50.82			
29		0.05068	50.83			
30		0.05046	50.61			

Measurement Uncertainty : \pm 0.12 μ l

Conclusion / Remarks:

- 1 Measurement uncertainty is at confidence level 95% which corresponds to a coverage factor of k= 2.23
- 2 Calibration is performed as per ISO 8655 - 6 : 2002 (E)
- 3 Gravimetric Method is adopted for calibration

Calibrated By

Janardhan S
 Janardhan S
 (Calibration Engineer)

Checked By

Chitra
 Chitrangadha P R
 (Sr. Calibration Engineer)



Authorised By

Shreyas B V
 Shreyas B V
 (Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Sarada Krishna Homoeopathic Medical College Hospital
Kulashekaram, Kanyakumari District, Tamil Nadu-629161

Customer's Reference :

SRF No. : TSC/18-19/7168

Dated : 12 Oct 2018

ULR.NO CC223118000004187F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/7168-4	22 Oct 2018	22 Apr 2019	1 of 2

Details of device under calibration

Transcal ID : TSC81206

Nomenclature : Micro Pipette	No. of Pages : 2
Make : Labopette	Cal Procedure No. : TSC/CAL/610
Model/Range : 500µl	DUC Received : 12 Oct 2018
SI No. : NA	DUC Condition on Receipt : Satisfactory
ID No. : SKHMC/BM/PPT-008	Cal At : Mechanical Lab

Environmental Conditions : Temperature in °C : 20.9

Humidity in RH % : 49

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Electronic Weighing Balance	Mettler Toledo	AG 285	1120102251	TSC/17-18/INH/MECH-19	10 Feb 2019

Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By

Janardhan S
(Calibration Engineer)

Checked By

Chitragadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas B V
(Lab Incharge)



Results:

Range : 500 μ l

ATM Pressure : 909.9 hpa

Sl. No.	Micropipette Set Volume in μ l	Standard Balance Reading in g	Actual Calculated Volume @ 27°C in μ l	Average Volume in μ l	Systematic Error, \pm in %	Random Error in \pm in %
1	500	0.49967	501.17	501.35	0.27	0.03
2		0.49970	501.20			
3		0.49976	501.26			
4		0.49981	501.31			
5		0.49989	501.39			
6		0.49993	501.43			
7		0.49997	501.47			
8		0.50005	501.55			
9		0.50012	501.62			
10		0.49965	501.15			

Measurement Uncertainty : \pm 0.20 μ l

Conclusion / Remarks:

- 1 Measurement uncertainty is at confidence level 95% which corresponds to a coverage factor of $k = 2.26$
- 2 Calibration is performed as per ISO 8655 - 6 : 2002 (E)
- 3 Gravimetric Method is adopted for calibration

Calibrated By

Janardhan S
(Calibration Engineer)

Checked By

Chitragadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas B V
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Sarada Krishna Homoeopathic Medical College Hospital
Kulashekaram, Kanyakumari District, Tamil Nadu-629161

Customer's Reference :

SRF No. : TSC/18-19/7168

Dated : 12 Oct 2018

ULR.NO CC223118000004188F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/7168-3	22 Oct 2018	22 Apr 2019	1 of 2

Details of device under calibration

Transcal ID : TSC106591

Nomenclature : Micro Pipette	No. of Pages : 2
Make : Astra	Cal Procedure No. : TSC/CAL/610
Model/Range : 5-50µl	DUC Received : 12 Oct 2018
SI No. : DY 21026	DUC Condition on Receipt : Satisfactory
ID No. : SKHMC/BM/PPT-002	Cal At : Mechanical Lab

Environmental Conditions : Temperature in °C : 20.8

Humidity in RH % : 49

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Electronic Weighing Balance	Mettler Toledo	AG 285	1120102251	TSC/17-18/INH/MECH-19	10 Feb 2019
2	Micro Balance	Sartorius	MC 21 S	16406121	TSC/17-18/INH/Mech-24	10 Mar 2019

Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By

Janardhan S
(Calibration Engineer)

Checked By

Chitrangadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas B V
(Lab Incharge)



Range : 5 - 50 μ l
 Increment : 0.1 μ l

ATM Pressure : 909.7 hpa

Sl. No.	Micropipette Set Volume in μ l	Standard Balance Reading in g	Actual Calculated Volume @ 27°C in μ l	Average Volume in μ l	Systematic Error, \pm in %	Random Error, in \pm in %
1	5	0.004928	4.943	4.960	-0.80	0.29
2		0.004931	4.946			
3		0.004937	4.952			
4		0.004942	4.957			
5		0.004949	4.964			
6		0.004952	4.967			
7		0.004955	4.970			
8		0.004964	4.979			
9		0.004968	4.983			
10		0.004928	4.943			
11	25	0.02517	25.25	25.36	1.42	0.36
12		0.02520	25.28			
13		0.02522	25.30			
14		0.02526	25.34			
15		0.02529	25.37			
16		0.02534	25.42			
17		0.02535	25.43			
18		0.02537	25.45			
19		0.02543	25.51			
20		0.02517	25.25			
21	50	0.05064	50.79	50.89	1.79	0.17
22		0.05067	50.82			
23		0.05068	50.83			
24		0.05072	50.87			
25		0.05076	50.91			
26		0.05079	50.94			
27		0.05080	50.95			
28		0.05083	50.98			
29		0.05089	51.04			
30		0.05064	50.79			

Measurement Uncertainty : \pm 0.12 μ l

Conclusion / Remarks:

- 1 Measurement uncertainty is at confidence level 95% which corresponds to a coverage factor of k= 2.23
- 2 Calibration is performed as per ISO 8655 - 6 : 2002 (E)
- 3 Gravimetric Method is adopted for calibration

Calibrated By

Janardhan S
 Janardhan S
 (Calibration Engineer)

Checked By

Chitra
 Chitragadha P R
 (Sr. Calibration Engineer)



Authorised By

Shreyas B V
 Shreyas B V
 (Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Sarada Krishna Homoeopathic Medical College Hospital
Kulashekaram, Kanyakumari District, Tamil Nadu-629161

Customer's Reference :

SRF No. : TSC/18-19/7168

Dated : 12 Oct 2018

ULR.NO CC22311800004183F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/7168-2	22 Oct 2018	22 Apr 2019	1 of 2

Details of device under calibration		Transcal ID : TSC46528	
Nomenclature	: Micro Pipette	No. of Pages	: 2
Make	: Astra	Cal Procedure No.	: TSC/CAL/610
Model/Range	: 500µl	DUC Received	: 12 Oct 2018
SI No.	: DX81325	DUC Condition on Receipt	: Satisfactory
ID No.	: SKHMC/BM/PPT-011	Cal At	: Mechanical Lab

Environmental Conditions : Temperature in °C : 20.7

Humidity in RH % : 49

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Electronic Weighing Balance	Mettler Toledo	AG 285	1120102251	TSC/17-18/INH/MECH-19	10 Feb 2019

Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By

Janardhan S
(Calibration Engineer)

Checked By

Chitra
Chitrangadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas B V
(Lab Incharge)



Results:

Range : 500 µl

ATM Pressure : 909.7 hpa

Sl. No.	Micropipette Set Volume in µl	Standard Balance Reading in g	Actual Calculated Volume @ 27°C in µl	Average Volume in µl	Systematic Error, ± in %	Random Error in ± in %
1	500	0.49959	501.09	501.22	0.24	0.02
2		0.49962	501.12			
3		0.49967	501.17			
4		0.49970	501.20			
5		0.49971	501.21			
6		0.49978	501.28			
7		0.49979	501.29			
8		0.49986	501.36			
9		0.49993	501.43			
10		0.49957	501.07			

Measurement Uncertainty : ± 0.20 µl

Conclusion / Remarks:

- 1 Measurement uncertainty is at confidence level 95% which corresponds to a coverage factor of k = 2.23
- 2 Calibration is performed as per ISO 8655 - 6 : 2002 (E)
- 3 Gravimetric Method is adopted for calibration

Calibrated By

Janardhan S
(Calibration Engineer)

Checked By

Chitragadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas B V
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Sarada Krishna Homoeopathic Medical College Hospital
Kulashekaram, Kanyakumari District, Tamil Nadu-629161

Customer's Reference : SRF No. : TSC/18-19/7168 Dated : 12 Oct 2018
ULR.NO CC22311800004181F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/7168-1	22 Oct 2018	22 Apr 2019	1 of 2

Details of device under calibration		Transcal ID : TSC46534	
Nomenclature	: Micro Pipette	No. of Pages	: 2
Make	: Dragon Lab	Cal Procedure No.	: TSC/CAL/610
Model/Range	: 1000µl	DUC Received	: 12 Oct 2018
SI No.	: YE5A481592	DUC Condition on Receipt	: Satisfactory
ID No.	: SKHMC/BM/PPT-007	Cal At	: Mechanical Lab

Environmental Conditions : Temperature in °C : 20.7 Humidity in RH % : 49

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Electronic Weighing Balance	Mettler Toledo	AG 285	1120102251	TSC/17-18/INH/MECH-19	10 Feb 2019

Note :

- This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
- Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
- Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
- Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
- Calibration of the DUC are traceable to National/International Standards
- Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
- In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
- Consider Model or Range whichever is applicable.
- Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By

Janardhan S
(Calibration Engineer)

Checked By

Chitragadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas B V
(Lab Incharge)



Results:

Range : 1000 μ l

ATM Pressure : 909.7 hpa

Sl. No.	Micropipette Set Volume in μ l	Standard Balance Reading in g	Actual Calculated Volume @ 27°C in μ l	Average Volume in μ l	Systematic Error, \pm in %	Random Error in \pm in %
1	1000	1.00007	1003.07	1003.23	0.32	0.01
2		1.00015	1003.15			
3		1.00021	1003.21			
4		1.00020	1003.20			
5		1.00026	1003.26			
6		1.00027	1003.27			
7		1.00030	1003.30			
8		1.00034	1003.34			
9		1.00042	1003.42			
10		1.00005	1003.05			

Measurement Uncertainty : \pm 0.20 μ l

Conclusion / Remarks:

- 1 Measurement uncertainty is at confidence level 95% which corresponds to a coverage factor of $k = 2.23$
- 2 Calibration is performed as per ISO 8655 - 6 : 2002 (E)
- 3 Gravimetric Method is adopted for calibration

Calibrated By

Janardhan S
(Calibration Engineer)

Checked By

Chitragadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas B V
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At :100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Sarada Krishna Homoeopathic Medical College Hospital
Kulashekaram, Kanyakumari District, Tamil Nadu-629161

Customer's Reference :

SRF No. : TSC/18-19/7168

Dated : 12 Oct 2018

ULR.NO CC223118000004193F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/7168-9	13 Oct 2018	13 Oct 2019	1 of 2

Details of device under calibration

Transcal ID

: TSC142341

Nomenclature : Height Scale	No. of Pages : 2
Make : Biocon	Cal Procedure No. : TSC/CAL/428
Model/Range : 0-200cm	DUC Received : 12 Oct 2018
SI No. :	DUC Condition on Receipt : Satisfactory
ID No. : SKHMC/BM/H S-002	Cal At : Dimensional Lab

Environmental Conditions : Temperature in °C : 20.1

Humidity in RH % : 52.0

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Length Measuring Machine	ME	0-1000 mm	010300	CGIL/2017/0422/C-001	21 Apr 2019

Note :

- This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
- Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
- Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
- Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
- Calibration of the DUC are traceable to National/International Standards
- Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
- In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
- Consider Model or Range whichever is applicable.
- Nabl-133 guidelines are adopted for use of NABL symbol.



Calibrated By

Anandesh T N
(Calibration Engineer)

Checked By

Pavan P K
(Calibration Engineer)

Authorised By

Anandmurthy B.S
(Lab Incharge)



CAL CERT. NO : TSC/18-19/7168-9

Page : 2 of 2

Range: 0 - 2000 mm

L.C.: 1 mm

Results:

Sl. No.	DUC Marking in mm	STD Readings in mm	Error Observed in mm
1	150	150.000	0.000
2	200	199.980	0.020
3	300	300.195	-0.195
4	500	500.640	-0.640
5	750	750.015	-0.015
6	1000	1001.220	-1.220
7	1500	1501.875	-1.875
8	2000	2002.835	-2.835

Conclusion/ Remarks:

Limits of Tolerance as per IS:1269-1997

Class	Maximum permissible error for intial verification
Class I	$\pm (0.1 + 0.1L)$ mm
Class II	$\pm (0.3 + 0.2L)$ mm
Class III	$\pm (0.6 + 0.4L)$ mm

"L" is length of the measuring tape in meter

Maximum permissible error for Measures in Service is Equal to twice the maximum permissible error at the the time of intial verification.


Measurement Uncertainty is $\pm (565 + L/100)$ μ m at 95% Confidence level , with k=1.96



Calibrated By


Anandesh T N
(Calibration Engineer)

Checked By


Pavan P K
(Calibration Engineer)

Authorised By


Anandmurthy B.S
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At :100, W.Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin- 560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Sarada Krishna Homoeopathic Medical College Hospital
Kulashekaram, Kanyakumari District, Tamil Nadu-629161

Customer's Reference :

SRF No. : TSC/18-19/7168

Dated : 12 Oct 2018

ULR.NO CC223118000004191F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/7168-8	13 Oct 2018	13 Oct 2019	1 of 2

Details of device under calibration		Transcal ID	
Nomenclature	: Height Scale	No. of Pages	: 2
Make	: Biocon	Cal Procedure No.	: TSC/CAL/428
Model/Range	: 0-200cm	DUC Received	: 12 Oct 2018
SI No.	:	DUC Condition on Receipt	: Satisfactory
ID No.	: SKHMC/BM/H S001	Cal At	: Dimensional Lab

Environmental Conditions : Temperature in °C : 20.1

Humidity in RH % : 52.0

Standards used :

SI No.	Nomenclature	Make	Model	SI No/ID.No.	Certificate No.	Validity
1	Length Measuring Machine	ME	0-1000 mm	010300	CGIL/2017/0422/C-001	21 Apr 2019

Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.



Calibrated By

Anandesh T N
(Calibration Engineer)

Checked By

Pavan P K
(Calibration Engineer)

Authorised By

Anandmurthy B.S
(Lab Incharge)



CAL CERT. NO : TSC/18-19/7168-8

Page : 2 of 2

Range: 0 - 2000 mm

L.C.: 1 mm

Results:

Sl. No.	DUC Marking in mm	STD Readings in mm	Error Observed in mm
1	150	150.000	0.000
2	200	200.035	-0.035
3	300	300.365	-0.365
4	500	501.030	-1.030
5	750	751.785	-1.785
6	1000	1001.595	-1.595
7	1500	1503.030	-3.030
8	2000	2004.245	-4.245

Conclusion/ Remarks:

Limits of Tolerance as per IS:1269-1997

Class	Maximum permissible error for initial verification
Class I	$\pm (0.1 + 0.1L)$ mm
Class II	$\pm (0.3 + 0.2L)$ mm
Class III	$\pm (0.6 + 0.4L)$ mm

"L" is length of the measuring tape in meter

Maximum permissible error for Measures in Service is Equal to twice the maximum permissible error at the the time of initial verification.

Measurement Uncertainty is $\pm (565 + L/100)$ μ m at 95% Confidence level , with $k=1.96$



Calibrated By


Anandesh T N
(Calibration Engineer)

Checked By


Pavan P K
(Calibration Engineer)

Authorised By


Anandmurthy B.S
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld.Premises : Centenary Building (G.Flr), Door No.At :100, W.Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. **SARADA KRISHNA HOMOEOPATHIC MEDICAL HOSPITAL**
Kulashekaram, Kanyakumari Distric, Tamil Nadu-629161

Customer's Reference : **SRF No. : TSC/18-19/6688** Dated : 27 Sep 2018
ULR.NO CC223118000002325F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/6688-7	04 Oct 2018	04 Apr 2019	1 of 2

Details of device under calibration		Transcal ID : TSC140375	
Nomenclature	: Micro Pipette	No. of Pages	: 2
Make	: Accuvet	Cal Procedure No.	: TSC/CAL/610
Model/Range	: 25µl	DUC Received	: 27 Sep 2018
SI No.	:	DUC Condition on Receipt	: Satisfactory
ID No.	: SKHMC/BM/PPT004	Cal At	: Mechanical Lab

Environmental Conditions : Temperature in °C : 20.9

Humidity in RH % : 49

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Electronic Weighing Balance	Mettler Toledo	AG 285	1120102251	TSC/17-18/INH/MECH-19	10 Feb 2019

Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any , in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings areout of specification limit & ' - ' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By

Koushik
Koushik Gijgal K R
(Calibration Engineer)

Checked By

Chitra
Chitrangadha P R
(Sr.Calibration Engineer)



Authorised By

Shreyas
Shreyas B V
(Lab Incharge)



CAL CERT. NO : TSC/18-19/6688-7

Page : 2 of 2

Results:

Range : 25 μ l

ATM Pressure : 909.7 hpa

Sl. No.	Micropipette Set Volume in μ l	Standard Balance Reading in g	Actual Calculated Volume @ 27°C in μ l	Average Volume in μ l	Systematic Error, \pm in %	Random Error in \pm in %
1	25	0.02467	24.74	24.86	-0.56	0.40
2		0.02471	24.78			
3		0.02473	24.80			
4		0.02474	24.81			
5		0.02479	24.86			
6		0.02482	24.89			
7		0.02488	24.95			
8		0.02490	24.97			
9		0.02495	25.02			
10		0.02467	24.74			

Measurement Uncertainty : \pm 0.15 μ l

Conclusion / Remarks:

- 1 Measurement uncertainty is at confidence level 95% which corresponds to a coverage factor of $k = 1.96$
- 2 Calibration is performed as per ISO 8655 - 6 : 2002 (E)
- 3 Gravimetric Method is adopted for calibration

Calibrated By

Koushik

Koushik Gijgal K R
(Calibration Engineer)

Checked By

Chitra

Chitrangadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas
Shreyas B V
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. SARADA KRISHNA HOMOEOPATHIC MEDICAL HOSPITAL
Kulashekaram, Kanyakumari Distric, Tamil Nadu-629161

Customer's Reference :

SRF No. : TSC/18-19/6688

Dated : 27 Sep 2018

ULR.NO CC223118000002326F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/6688-6	04 Oct 2018	04 Apr 2019	1 of 2

Details of device under calibration

Transcal ID : TSC140374

Nomenclature : Micro Pipette	No. of Pages : 2
Make : Dragon	Cal Procedure No. : TSC/CAL/610
Model/Range : 5-50µl	DUC Received : 27 Sep 2018
SI No. : YE16CAA0129851	DUC Condition on Receipt : Satisfactory
ID No. : --	Cal At : Mechanical Lab

Environmental Conditions : Temperature in °C : 20.9

Humidity in RH % : 49

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Electronic Weighing Balance	Mettler Toledo	AG 285	1120102251	TSC/17-18/INH/MECH-19	10 Feb 2019
2	Micro Balance	Sartorius	MC 21 S	16406121	TSC/17-18/INH/Mech-24	10 Mar 2019

Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By

Koushik

Koushik Gijgal K R
(Calibration Engineer)

Checked By

Chitra

Chitrangadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas
Shreyas B V
(Lab Incharge)



CAL CERT. NO : TSC/18-19/6688-6

Page : 2 of 2

Range : 5 - 50 μ l
 Increment : 0.5 μ l

ATM Pressure : 909.6 hpa

Sl. No.	Micropipette Set Volume in μ l	Standard Balance Reading in g	Actual Calculated Volume @ 27°C in μ l	Average Volume in μ l	Systematic Error, \pm in %	Random Error, in \pm in %
1	5	0.004768	4.782	4.828	-3.43	0.78
2		0.004782	4.796			
3		0.004789	4.803			
4		0.004794	4.808			
5		0.004832	4.846			
6		0.004838	4.852			
7		0.004844	4.859			
8		0.004859	4.874			
9		0.004865	4.880			
10		0.004768	4.782			
11	25	0.02502	25.10	25.23	0.92	0.42
12		0.02507	25.15			
13		0.02511	25.19			
14		0.02513	25.21			
15		0.02518	25.26			
16		0.02520	25.28			
17		0.02522	25.30			
18		0.02526	25.34			
19		0.02534	25.42			
20		0.02502	25.10			
21	50	0.05078	50.93	51.06	2.11	0.20
22		0.05080	50.95			
23		0.05083	50.98			
24		0.05089	51.04			
25		0.05093	51.08			
26		0.05097	51.12			
27		0.05098	51.13			
28		0.05102	51.17			
29		0.05105	51.20			
30		0.05078	50.93			

Measurement Uncertainty : \pm 0.12 μ l

Conclusion / Remarks:

- 1 Measurement uncertainty is at confidence level 95% which corresponds to a coverage factor of $k=2.26$
- 2 Calibration is performed as per ISO 8655 - 6 : 2002 (E)
- 3 Gravimetric Method is adopted for calibration

Calibrated By

Koushik

Koushik Gijgal K R
(Calibration Engineer)

Checked By

Chitra

Chitrangadha P R
(Sr. Calibration Engineer)



Authorized By

Shreyas
Shreyas B Y
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At :100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. SARADA KRISHNA HOMOEOPATHIC MEDICAL HOSPITAL
Kulashekaram, Kanyakumari Distric, Tamil Nadu-629161

Customer's Reference :

SRF No. : TSC/18-19/6688

Dated : 27 Sep 2018

ULR.NO CC223118000002322F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/6688-5	04 Oct 2018	04 Apr 2019	1 of 2

Details of device under calibration

Transcal ID : TSC46503

Nomenclature : Micro Pipette	No. of Pages : 2
Make : Astra	Cal Procedure No. : TSC/CAL/610
Model/Range : 10µl	DUC Received : 27 Sep 2018
SI No. : DV27600	DUC Condition on Receipt : Satisfactory
ID No. : SKHMC/BM/PPT-012	Cal At : Mechanical Lab

Environmental Conditions : Temperature in °C : 20.9

Humidity in RH % : 49

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Micro Balance	Sartorius	MC 21 S	16406121	TSC/17-18/INH/Mech-24	10 Mar 2019

Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By

Koushik

Koushik Gijgal K R
(Calibration Engineer)

Checked By

Chitra

Chitrangadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas
Shreyas B V
(Lab Incharge)



Results:

Range : 10 μ l

ATM Pressure : 909.7 hpa

Sl. No.	Micropipette Set Volume in μ l	Standard Balance Reading in g	Actual Calculated Volume @ 27°C in μ l	Average Volume in μ l	Systematic Error, \pm in %	Random Error in \pm in %
1	10	0.009872	9.902	9.927	-0.73	0.22
2		0.009879	9.909			
3		0.009885	9.915			
4		0.009890	9.920			
5		0.009896	9.926			
6		0.009911	9.941			
7		0.009917	9.947			
8		0.009924	9.954			
9		0.009929	9.959			
10		0.009872	9.902			

Measurement Uncertainty : \pm 0.03 μ l

Conclusion / Remarks:

- 1 Measurement uncertainty is at confidence level 95% which corresponds to a coverage factor of $k = 1.96$
- 2 Calibration is performed as per ISO 8655 - 6 : 2002 (E)
- 3 Gravimetric Method is adopted for calibration

Calibrated By

Koushik

Koushik Gijgal K R
(Calibration Engineer)

Checked By

Chitra

Chitragadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas
Shreyas B V
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises: Centenary Building (G. Flr), Door No. At :100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. SARADA KRISHNA HOMOEOPATHIC MEDICAL HOSPITAL
Kulashekaram, Kanyakumari Distric, Tamil Nadu-629161

Customer's Reference :

SRF No. : TSC/18-19/6688

Dated : 27 Sep 2018

ULR.NO CC223118000002323F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/6688-4	04 Oct 2018	04 Apr 2019	1 of 2

Details of device under calibration

Transcal ID

: TSC140378

Nomenclature : Micro Pipette	No. of Pages : 2
Make : Ozopet	Cal Procedure No. : TSC/CAL/610
Model/Range : 50µl	DUC Received : 27 Sep 2018
SI No. : NA	DUC Condition on Receipt : Satisfactory
ID No. : --	Cal At : Mechanical Lab

Environmental Conditions : Temperature in °C : 21.0

Humidity in RH % : 49

Standards used :

SI No.	Nomenclature	Make	Model	SI No/ID.No.	Certificate No.	Validity
1	Electronic Weighing Balance	Mettler Toledo	AG 285	1120102251	TSC/17-18/INH/MECH-19	10 Feb 2019

Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any , in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By

Koushik

Koushik Gijgal K R
(Calibration Engineer)

Checked By

Chitra

Chitrangadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas
Shreyas B V
(Lab Incharge)



Results:

Range : 50 µl

ATM Pressure : 909.6 hpa

Sl. No.	Micropipette Set Volume in µl	Standard Balance Reading in g	Actual Calculated Volume @ 27°C in µl	Average Volume in µl	Systematic Error, ± in %	Random Error in ± in %
1	50	0.05010	50.25	50.40	0.80	0.25
2		0.05012	50.27			
3		0.05017	50.32			
4		0.05024	50.39			
5		0.05026	50.41			
6		0.05031	50.46			
7		0.05035	50.50			
8		0.05039	50.54			
9		0.05044	50.59			
10		0.05010	50.25			

Measurement Uncertainty : ± 0.16 µl

Conclusion / Remarks:

- 1 Measurement uncertainty is at confidence level 95% which corresponds to a coverage factor of k = 1.96
- 2 Calibration is performed as per ISO 8655 - 6 : 2002 (E)
- 3 Gravimetric Method is adopted for calibration

Calibrated By

Koushik

Koushik Gijgal K R
(Calibration Engineer)

Checked By

Chitra

Chitrangadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas B
Shreyas B
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. SARADA KRISHNA HOMOEOPATHIC MEDICAL HOSPITAL
Kulashekaram, Kanyakumari Distric, Tamil Nadu-629161

Customer's Reference : SRF No. : TSC/18-19/6688 Dated : 27 Sep 2018
ULR.NO CC223118000002327F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/6688-2	04 Oct 2018	04 Apr 2019	1 of 2

Details of device under calibration		Transcal ID	
Nomenclature	: Micro Pipette	Transcal ID	: TSC108499
Make	: Thermo Scientific	No. of Pages	: 2
Model/Range	: 100-1000µl	Cal Procedure No.	: TSC/CAL/610
SI No.	: NW07420	DUC Received	: 27 Sep 2018
ID No.	: --	DUC Condition on Receipt	: Satisfactory
		Cal At	: Mechanical Lab

Environmental Conditions : Temperature in °C : 20.9 Humidity in RH % : 49

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Electronic Weighing Balance	Mettler Toledo	AG 285	1120102251	TSC/17-18/INH/MECH-19	10 Feb 2019

Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any , in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By

Koushik

Koushik Gijgal K R
(Calibration Engineer)

Checked By

Chitra

Chitrangadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas

Shreyas B V
(Lab Incharge)



Range : 100-1000 ml
 Increment : 1

ATM Pressure : 909.7 hpa

Sl. No.	Micropipette Set Volume in μ l	Standard Balance Reading in g	Actual Calculated Volume @ 27°C in μ l	Average Volume in μ l	Systematic Error, \pm in %	Random Error, in \pm in %
1	100	0.09983	100.13	100.29	0.29	0.15
2		0.09989	100.19			
3		0.09990	100.20			
4		0.09995	100.25			
5		0.09997	100.27			
6		0.10002	100.32			
7		0.10012	100.42			
8		0.10019	100.49			
9		0.10024	100.54			
10		0.09981	100.11			
11	500	0.50553	507.05	507.26	1.45	0.04
12		0.50559	507.11			
13		0.50564	507.16			
14		0.50572	507.24			
15		0.50577	507.29			
16		0.50580	507.32			
17		0.50589	507.41			
18		0.50593	507.45			
19		0.50606	507.58			
20		0.50551	507.03			
21	1000	1.00452	1007.53	1007.70	0.77	0.02
22		1.00455	1007.56			
23		1.00459	1007.60			
24		1.00463	1007.64			
25		1.00467	1007.68			
26		1.00470	1007.71			
27		1.00479	1007.80			
28		1.00492	1007.93			
29		1.00498	1007.99			
30		1.00450	1007.51			

Measurement Uncertainty : \pm 0.20

Conclusion / Remarks:

- 1 Measurement uncertainty is at confidence level 95% which corresponds to a coverage factor of $k=2.26$
- 2 Calibration is performed as per ISO 8655 - 6 : 2002 (E)
- 3 Gravimetric Method is adopted for calibration

Calibrated By

Koushik

Koushik Gijgal K R
(Calibration Engineer)

Checked By

Chitra

Chitrangadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas B
Shreyas B
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At :100, W.Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. SARADA KRISHNA HOMOEOPATHIC MEDICAL HOSPITAL
Kulashekaram, Kanyakumari Distric, Tamil Nadu-629161

Customer's Reference : SRF No. : TSC/18-19/6688 Dated : 27 Sep 2018

ULR.NO CC223118000002321F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/6688-1	04 Oct 2018	04 Apr 2019	1 of 2

Details of device under calibration		Transcal ID	
Nomenclature	: Micro Pipette	No. of Pages	: 2
Make	: Thermo Scientific	Cal Procedure No.	: TSC/CAL/610
Model/Range	: 100-1000µl	DUC Received	: 27 Sep 2018
SI No.	: NW 07418	DUC Condition on Receipt	: Satisfactory
ID No.	: --	Cal At	: Mechanical Lab

Environmental Conditions : Temperature in °C : 20.9

Humidity in RH % : 49

Standards used :

SI No.	Nomenclature	Make	Model	SI No/ID.No.	Certificate No.	Validity
1	Electronic Weighing Balance	Mettler Toledo	AG 285	1120102251	TSC/17-18/INH/MECH-19	10 Feb 2019

Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any , in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By

Koushik

Koushik Gijgal K R
(Calibration Engineer)

Checked By

Chitra

Chitrangadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas
Shreyas B V
(Lab Incharge)



Range : 100-1000 µl
 Increment : 1 µl

ATM Pressure : 909.7 hpa

Sl. No.	Micropipette Set Volume in µl	Standard Balance Reading in g	Actual Calculated Volume @ 27°C in µl	Average Volume in µl	Systematic Error, ± in %	Random Error, in ± in %
1	100	0.09986	100.16	100.31	0.31	0.13
2		0.09989	100.19			
3		0.09994	100.24			
4		0.09999	100.29			
5		0.10003	100.33			
6		0.10010	100.40			
7		0.10014	100.44			
8		0.10015	100.45			
9		0.10021	100.51			
10		0.09984	100.14			
11	500	0.50682	508.34	508.56	1.71	0.03
12		0.50695	508.47			
13		0.50699	508.51			
14		0.50705	508.57			
15		0.50706	508.58			
16		0.50712	508.64			
17		0.50717	508.69			
18		0.50720	508.72			
19		0.50724	508.76			
20		0.50680	508.32			
21	1000	1.01067	1013.70	1013.85	1.38	0.01
22		1.01072	1013.75			
23		1.01077	1013.80			
24		1.01078	1013.81			
25		1.01085	1013.88			
26		1.01088	1013.91			
27		1.01090	1013.93			
28		1.01096	1013.99			
29		1.01099	1014.02			
30		1.01065	1013.68			

Measurement Uncertainty : ± 0.20 µl

Conclusion / Remarks:

- 1 Measurement uncertainty is at confidence level 95% which corresponds to a coverage factor of k= 2.26
- 2 Calibration is performed as per ISO 8655 - 6 : 2002 (E)
- 3 Gravimetric Method is adopted for calibration

Calibrated By

Koushik

Koushik Gijgal K R
(Calibration Engineer)

Checked By

Chitra

Chitrangadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas

Shreyas B V
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. SARADA KRISHNA HOMOEOPATHIC MEDICAL HOSPITAL
Kulashekaram, Kanyakumari Distric, Tamil Nadu-629161

Customer's Reference : SRF No. : TSC/18-19/6688 Dated : 27 Sep 2018
ULR.NO CC223118000002324F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/6688-8	04 Oct 2018	04 Apr 2019	1 of 2

Details of device under calibration		Transcal ID : TSC140377	
Nomenclature	: Micro Pipette	No. of Pages	: 2
Make	: --	Cal Procedure No.	: TSC/CAL/610
Model/Range	: 100µl	DUC Received	: 27 Sep 2018
SI No.	: NA	DUC Condition on Receipt	: Satisfactory
ID No.	: SKHMC/BM/PPT005	Cal At	: Mechanical Lab

Environmental Conditions : Temperature in °C : 20.8 Humidity in RH % : 49

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Electronic Weighing Balance	Mettler Toledo	AG 285	1120102251	TSC/17-18/INH/MECH-19	10 Feb 2019

Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any , in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By

Koushik

Koushik Gijgal K R
(Calibration Engineer)

Checked By

Chitra

Chitrangadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas
Shreyas B V
(Lab Incharge)



CAL CERT. NO : TSC/18-19/6688-8

Page : 2 of 2

Results:

Range : 100 μ l

ATM Pressure : 909.7 hpa

Sl. No.	Micropipette Set Volume in μ l	Standard Balance Reading in g	Actual Calculated Volume @ 27°C in μ l	Average Volume in μ l	Systematic Error, \pm in %	Random Error in \pm in %
1	100	0.09975	100.05	100.20	0.20	0.12
2		0.09978	100.08			
3		0.09981	100.11			
4		0.09989	100.19			
5		0.09993	100.23			
6		0.09996	100.26			
7		0.09997	100.27			
8		0.10003	100.33			
9		0.10009	100.39			
10		0.09975	100.05			

Measurement Uncertainty : \pm 0.12 μ l

Conclusion / Remarks:

- 1 Measurement uncertainty is at confidence level 95% which corresponds to a coverage factor of $k = 2.23$
- 2 Calibration is performed as per ISO 8655 - 6 : 2002 (E)
- 3 Gravimetric Method is adopted for calibration

Calibrated By

Koushik
Koushik Gijgal K R
(Calibration Engineer)

Checked By

Chitra
Chitragadha P R
(Sr. Calibration Engineer)



Authorised By

Shreyas
Shreyas B V
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Address : M/s. Sarada Krishna Homeopathic Medical College,
Kanyakumari District, Kulasekaram,

Customer's Reference : SRF No.: 204 Date : 04 Sep 2018

CAL CERT. NO.	CALIBRATED ON	RECOMMENDED CAL DUE	PAGE NO.
TSC / 18 - 19 / 204 07	04 Sep 2018	04 Sep 2019	1 of 2

Details of device under calibration (DUC) :

DUC : Refrigerator	Cal Procedure No. : TSC/CAL/509
Make : Godrej	No. of Pages : 02
Model : RD Edge 185 E2H	DUC Received : 04 Sep 2018
Sl. No. : 1610033998	DUC Condition on Receipt : Satisfactory
Asset No. : NA	Location : Laboratory

Environmental Conditions : Temperature: $25 \pm 3^{\circ}\text{C}$ Humidity: 45 - 75% RH

Standards used :

Sl. No.	Nomenclature	Make & Model	Sl. No.	Uncertainty / Basic Accuracy (\pm)	Traceable to & Cert. No.	Validity
1	Paperless Recorder	Brain Child &VR 18	VR 009079	0.12°C	TransCal & TSC/17-18/INH/TCAL/58	16 Dec 2018
2	RTD Sensor	PT 100	RTD/011 to 015	0.13°C	TransCal/ TSC/17-18/INH/TCAL/78A to 78E	13 Nov 2018

Note :

1. The Calibration Certificate relates only to the above DUC.
2. Publication or reproduction of this cert., in any form other than by complete set of the whole certificate & in the language, written, is not permitted without the written consent of Transcal.
3. Corrections/erasing, invalidate the Calibration Certificate.
4. Calibration of the DUC are traceable to National standards/International Standards.
5. Any error in this certificate should be brought to our knowledge within 45 days from the date of this Certificate.
6. Results Reported are valid at the time of and under the stated conditions of measurements.
7. The usage of NABL Symbol is as per NABL guidelines given in NABL-133.

Calibrated By

Jitesh

(Calibration Engineer)

Checked By

Naveen Kumar S R

(Sr. Calibration Engineer)



Authorised By

Bhanuprakash G M

(Onsite Manager)



CAL CERT. NO.: TSC / 18 - 19 / 204 07

Page No. : 2 of 2

Results :-

Temperature Verification :-

Sl. No.	DUC Setting (°C)	DUC Reading (°C)	Standard Reading in °C					Uniformity in °C
			1	2	3	4	5	
1	5.6	5.7	5.6	5.7	5.5	5.8	5.9	0.4

Note : Expanded Uncertainty is $\pm 0.78^{\circ}\text{C}$.

Conclusion / Remarks :-

1. Measurement Uncertainty reported is at approximately 95% confidence level with $k=2$.

Calibrated By

Jitesh

(Calibration Engineer)

Checked By

Naveen Kumar S R
(Sr. Calibration Engineer)



Authorised By

Bhanuprakash G M
(Onsite Manager)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin- 560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Address : M/s. Sarada Krishna Homeopathic Medical College,
Kanyakumari District, Kulasekaram,

Customer's Reference : SRF No.: 204 **Date :** 04 Sep 2018

CAL CERT. NO.	CALIBRATED ON	RECOMMENDED CAL DUE	PAGE NO.
TSC / 18 - 19 / 204 08	04 Sep 2018	04 Sep 2019	1 of 2

Details of device under calibration (DUC) :

DUC : Dry Bath	Cal Procedure No. : TSC/CAL/509
Make : Labtech Medico	No. of Pages : 02
Model : DB-11	DUC Received : 04 Sep 2018
Sl. No. : 3532	DUC Condition on Receipt : Satisfactory
Asset No. : NA	Location : Laboratory

Environmental Conditions : Temperature: 25 ± 3° C Humidity: 45 - 75% RH

Standards used :

Sl. No.	Nomenclature	Make & Model	Sl. No.	Uncertainty / Basic Accuracy (±)	Traceable to & Cert. No.	Validity
1	Paperless Recorder	Brain Child & VR 18	VR 009079	0.12°C	TransCal & TSC/17-18/INH/TCAL/58	16 Dec 2018
2	RTD Sensor	PT 100	RTD/011 to 015	0.13°C	TransCal/ TSC/17-18/INH/TCAL/78A to 78E	13 Nov 2018

Note :

- The Calibration Certificate relates only to the above DUC.
- Publication or reproduction of this cert., in any form other than by complete set of the whole certificate & in the language, written, is not permitted without the written consent of Transcal.
- Corrections/erasing, invalidate the Calibration Certificate.
- Calibration of the DUC are traceable to National standards/International Standards.
- Any error in this certificate should be brought to our knowledge within 45 days from the date of this Certificate.
- Results Reported are valid at the time of and under the stated conditions of measurements.
- The usage of NABL Symbol is as per NABL guidelines given in NABL-133.

Calibrated By


Jitesh

(Calibration Engineer)

Checked By



Naveen Kumar S R
(Sr. Calibration Engineer)



Authorised By



Bhanuprakash G M
(Onsite Manager)



CAL CERT. NO.: TSC / 18 - 19 / 204 08

Page No. : 2 of 2

Results :-

Temperature Verification :-

Used Range : 36°C

Sl. No.	DUC Setting (°C)	DUC Reading (°C)	Standard Reading in °C					Uniformity in °C
			1	2	3	4	5	
1	36.5	36.7	36.7	36.6	36.8	36.5	36.2	0.6

Note : Expanded Uncertainty is $\pm 0.95^\circ\text{C}$.

Conclusion / Remarks :-

1. Measurement Uncertainty reported is at approximately 95% confidence level with $k=2$.

Calibrated By

Jitesh

(Calibration Engineer)

Checked By

Naveen Kumar S R
(Sr. Calibration Engineer)



Authorised By

Bhanuprakash G M
(Onsite Manager)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Address : M/s. Sarada Krishna Homeopathic Medical College,
Kanyakumari District, Kulasekaram,

Customer's Reference : SRF No.: 204 Date : 04 Sep 2018

CAL CERT. NO.	CALIBRATED ON	RECOMMENDED CAL DUE	PAGE NO.
TSC / 18 - 19 / 204 06	04 Sep 2018	04 Sep 2019	1 of 2

Details of device under calibration (DUC) :

DUC : Refrigerator	Cal Procedure No. : TSC/CAL/509
Make : Godrej	No. of Pages : 02
Model : Classic	DUC Received : 04 Sep 2018
Sl. No. : NA	DUC Condition on Receipt : Satisfactory
Asset No. : NA	Location : Laboratory

Environmental Conditions : Temperature: 25 ± 3° C Humidity: 45 - 75% RH

Standards used :

Sl. No.	Nomenclature	Make & Model	Sl. No.	Uncertainty / Basic Accuracy (±)	Traceable to & Cert. No.	Validity
1	Paperless Recorder	Brain Child & VR 18	VR 009079	0.12°C	TransCal & TSC/17-18/INH/TCAL/58	16 Dec 2018
2	RTD Sensor	PT 100	RTD/011 to 015	0.13°C	TransCal/ TSC/17-18/INH/TCAL/78A to 78E	13 Nov 2018

Note :

1. The Calibration Certificate relates only to the above DUC.
2. Publication or reproduction of this cert., in any form other than by complete set of the whole certificate & in the language, written, is not permitted without the written consent of Transcal.
3. Corrections/erasing, invalidate the Calibration Certificate.
4. Calibration of the DUC are traceable to National standards/International Standards.
5. Any error in this certificate should be brought to our knowledge within 45 days from the date of this Certificate.
6. Results Reported are valid at the time of and under the stated conditions of measurements.
7. The usage of NABL Symbol is as per NABL guidelines given in NABL-133.

Calibrated By

Jitesh

(Calibration Engineer)

Checked By

Naveen Kumar S R
(Sr. Calibration Engineer)



Authorised By

Bhanuprakash G M
(Onsite Manager)



CAL CERT. NO.: TSC / 18 - 19 / 204 06

Page No. : 2 of 2

Results :-

Temperature Verification :-

Sl. No.	DUC Setting (°C)	DUC Reading (°C)	Standard Reading in °C					Uniformity in °C
			1	2	3	4	5	
1	6.8	6.7	6.6	6.7	6.5	6.8	7.0	0.5

Note : Expanded Uncertainty is $\pm 0.86^{\circ}\text{C}$.

Conclusion / Remarks :-

1. Measurement Uncertainty reported is at approximately 95% confidence level with $k=2$.

Calibrated By

Jitesh
(Calibration Engineer)

Checked By

Naveen Kumar S R
(Sr. Calibration Engineer)



Authorised By

Bhanuprakash G M
(Onsite Manager)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At :100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin- 560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Address : M/s. Sarada Krishna Homeopathic Medical College,
Kanyakumari District, Kulasekaram,

Customer's Reference : SRF No.: 204 Date : 04 Sep 2018

CAL CERT. NO.	CALIBRATED ON	RECOMMENDED CAL DUE	PAGE NO.
TSC / 18 - 19 / 204 05	04 Sep 2018	04 Sep 2019	1 of 2

Details of device under calibration (DUC) :

DUC : Incubator	Cal Procedure No. : TSC/CAL/509
Make : Anton	No. of Pages : 02
Model : NA	DUC Received : 04 Sep 2018
Sl. No. : 9871099	DUC Condition on Receipt : Satisfactory
Asset No. : NA	Location : Laboratory

Environmental Conditions : Temperature: 25 ± 3° C Humidity: 45 - 75% RH

Standards used :

Sl. No.	Nomenclature	Make & Model	Sl. No.	Uncertainty / Basic Accuracy (±)	Traceable to & Cert. No.	Validity
1	Paperless Recorder	Brain Child & VR 18	VR 009079	0.12°C	TransCal & TSC/17-18/INH/TCAL/58	16 Dec 2018
2	RTD Sensor	PT 100	RTD/011 to 015	0.13°C	TransCal/ TSC/17-18/INH/TCAL/78A to 78E	13 Nov 2018

Note :


- The Calibration Certificate relates only to the above DUC.
- Publication or reproduction of this cert., in any form other than by complete set of the whole certificate & in the language, written, is not permitted without the written consent of Transcal.
- Corrections/erasing, invalidate the Calibration Certificate.
- Calibration of the DUC are traceable to National standards/International Standards.
- Any error in this certificate should be brought to our knowledge within 45 days from the date of this Certificate.
- Results Reported are valid at the time of and under the stated conditions of measurements.
- The usage of NABL Symbol is as per NABL guidelines given in NABL-133.

Calibrated By


Jitesh

(Calibration Engineer)

Checked By



Naveen Kumar S R
(Sr. Calibration Engineer)



Authorised By



Bhanuprakash G M
(Onsite Manager)



CAL CERT. NO.: TSC / 18 - 19 / 204 05

Page No. : 2 of 2

Results :-

Temperature Verification :-

Used Range : 36°C

Sl. No.	DUC Setting (°C)	DUC Reading (°C)	Standard Reading in °C					Uniformity in °C
			1	2	3	4	5	
1	36.3	36.5	36.6	36.4	36.5	36.7	36.3	0.4

Note : Expanded Uncertainty is $\pm 0.78^\circ\text{C}$.

Conclusion / Remarks :-

1. Measurement Uncertainty reported is at approximately 95% confidence level with $k=2$.

Calibrated By

Jitesh

(Calibration Engineer)

Checked By

Naveen Kumar S R
(Sr. Calibration Engineer)



Authorised By

Bhanuprakash G M
(Onsite Manager)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At :100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Address : M/s. Sarada Krishna Homeopathic Medical College,
Kanyakumari District, Kulasekaram,

Customer's Reference : SRF No.: 204 **Date :** 04 Sep 2018

CAL CERT. NO.	CALIBRATED ON	RECOMMENDED CAL DUE	PAGE NO.
TSC / 18 - 19 / 204 04	04 Sep 2018	04 Sep 2019	1 of 2

Details of device under calibration (DUC) :

DUC : Hot Air Oven	Cal Procedure No. : TSC/CAL/509
Make : Anton	No. of Pages : 02
Model : NA	DUC Received : 04 Sep 2018
Sl. No. : 965105	DUC Condition on Receipt : Satisfactory
Asset No. : NA	Location : Laboratory

Environmental Conditions : Temperature: 25 ± 3° C Humidity: 45 - 75% RH

Standards used :

Sl. No.	Nomenclature	Make & Model	Sl. No.	Uncertainty / Basic Accuracy (±)	Traceable to & Cert. No.	Validity
1	Paperless Recorder	Brain Child &VR 18	VR 009079	0.12°C	TransCal & TSC/17-18/INH/TCAL/58	16 Dec 2018
2	RTD Sensor	PT 100	RTD/011 to 015	0.13°C	TransCal/ TSC/17-18/INH/TCAL/78A to 78E	13 Nov 2018

Note :

- The Calibration Certificate relates only to the above DUC.
- Publication or reproduction of this cert., in any form other than by complete set of the whole certificate & in the language, written, is not permitted without the written consent of Transcal.
- Corrections/erasing, invalidate the Calibration Certificate.
- Calibration of the DUC are traceable to National standards/International Standards.
- Any error in this certificate should be brought to our knowledge within 45 days from the date of this Certificate.
- Results Reported are valid at the time of and under the stated conditions of measurements.
- The usage of NABL Symbol is as per NABL guidelines given in NABL-133.

Calibrated By

Jitesh

(Calibration Engineer)

Checked By

Naveen Kumar S R


(Sr. Calibration Engineer)



Authorised By

Bhanuprakash G M

(Onsite Manager)



CAL CERT. NO.: TSC / 18 - 19 / 204 04

Page No. : 2 of 2

Results :-

Temperature Verification :-

Used Range : 48°C

Sl. No.	DUC Setting (°C)	DUC Reading (°C)	Standard Reading in °C					Uniformity in °C
			1	2	3	4	5	
1	48.0	47.8	47.7	47.9	48.0	48.2	48.1	0.5

Note : Expanded Uncertainty is $\pm 0.86^\circ\text{C}$.

Conclusion / Remarks :-

1. Measurement Uncertainty reported is at approximately 95% confidence level with $k=2$.

Calibrated By

Jitesh

(Calibration Engineer)

Checked By

Naveen Kumar S R
(Sr. Calibration Engineer)



Authorised By

Bhanuprakash G M
(Onsite Manager)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Sarada Krishna Homeopathic Medical College.,
Kanyakumari Dist, Kulasekharam, Tamil Nadu IND 629161

Customer's Reference : SRF No. : TSC/18-19/204 Dated : 04 Sep 2018
ULR.NO CC22311800000645F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/204-1	04 Sep 2018	04 Mar 2019	1 of 2

Details of device under calibration		Transcal ID : TSC48789	
Nomenclature	: Centrifuge	No. of Pages	: 2
Make	: REMI	Cal Procedure No.	: TSC/CAL/611
Model/Range	: R-8C	DUC Received	: 04 Sep 2018
SI No.	: EBLC-4514	DUC Condition on Receipt	: Satisfactory
ID No.	: SKHMC/BM/CF001	Cal At	: Mechanical onsite

Environmental Conditions : Temperature in °C : 23.4 Humidity in RH % : 49

Standards used :

SI No.	Nomenclature	Make	Model	SI No/ID.No.	Certificate No.	Validity
1	Digital Tachometer	Lutron	DT 2236	L - 510221	SANS/54486/18	14 Mar 2019
2	Digital stop watch	Racer	--	'---	TSC/18-19/INH/Mech-74	15 May 2019


Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data (To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By



Tejas
(Calibration Engineer)

Checked By


Annapoorna M P
(Calibration Engineer)



Authorised By


Shreyas B
(Lab Incharge)



Results :-

RPM :

Range: 0-5250 rpm

Location: Laboratory

Sl. No.	DUC Setting in rpm	STD Reading in rpm	Observed Deviation in rpm	Measurement Uncertainty \pm in rpm
1	1000	989.9	10.1	2.13
2	2000	1985	15	2.00
3	3000	2979	21	3.00
4	4000	3991	9	4.00
5	5000	4995	5	5.00

Timer :

Range: 0-60 Min

Sl. No.	DUC Reading in Min	Standard Reading in min : sec : msec	Observed Deviation in min : sec : msec	Measurement Uncertainty \pm in Sec
1	1	01 : 00 : 02	- 00 : 00 : 02	1.2
2	3	02 : 59 : 75	00 : 00 : 25	3.6
3	5	04 : 59 : 60	00 : 00 : 40	6.0


Conclusion Remarks :-

1. Measurement uncertainty reported is at 95% confidence level.

Calibrated By



Tejas
(Calibration Engineer)

Checked By


Annapoorna M P
(Calibration Engineer)



Authorised By


Shreyas E V
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises: Centenary Building (G. Flr), Door No. At :100, W.Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Sarada Krishna Homeopathic Medical College.,
Kanyakumari Dist, Kulasekharam, Tamil Nadu IND 629161

Customer's Reference : SRF No. : TSC/18-19/204 Dated : 04 Sep 2018

ULR.NO CC22311800000649F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/204-2	04 Sep 2018	04 Mar 2019	1 of 2

Details of device under calibration		Transcal ID	
Nomenclature	: Centrifuge	Transcal ID	: TSC140602
Make	: REMI	No. of Pages	: 2
Model/Range	: C852	Cal Procedure No.	: TSC/CAL/611
SI No.	: KALC-4528	DUC Received	: 04 Sep 2018
ID No.	: SKHMC/BM/CF002	DUC Condition on Receipt	: Satisfactory
		Cal At	: Mechanical onsite

Environmental Conditions : Temperature in °C : 23.4

Humidity in RH % : 49

Standards used :

SI No.	Nomenclature	Make	Model	SI No/ID.No.	Certificate No.	Validity
1	Digital Tachometer	Lutron	DT 2236	L - 510221	SANS/54486/18	14 Mar 2019


Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part of this Report- provided for incorporation of additional data (To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By



Tejas
(Calibration Engineer)

Checked By


Annapoorna M P
(Calibration Engineer)



Authorised By


Shreyas B V
(Lab Incharge)

Results :-


Department: Laboratory

Sl. No.	DUC Set At	STD Reading in rpm	Measurement Uncertainty \pm in rpm
1	Level 1	788.6	0.58
2	Level 2	1255	1.60
3	Level 3	1796	2.83
4	Level 4	2531	4.08
5	Level 5	3149	5.34

Conclusion Remarks :-

1. Measurement uncertainty reported is at 95% confidence level.

Calibrated By



Tejas
(Calibration Engineer)

Checked By


Annapoorna M P
(Calibration Engineer)



Authorised By


Shreyas B V
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin - 560003

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Sarada Krishna Homeopathic Medical College.,
Kanyakumari Dist, Kulasekharam, Tamil Nadu IND 629161

Customer's Reference : SRF No. : TSC/18-19/204 Dated : 04 Sep 2018

ULR.NO CC223118000000650F

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/204-3	04 Sep 2018	04 Sep 2019	1 of 2

Details of device under calibration		Transcal ID	:	TSC83642
Nomenclature	: Rotary Shaker	No. of Pages	:	2
Make	: --	Cal Procedure No.	:	TSC/CAL/611
Model/Range	: --	DUC Received	:	04 Sep 2018
SI No.	: 3167	DUC Condition on Receipt	:	Satisfactory
ID No.	:	Cal At	:	Mechanical onsite

Environmental Conditions : Temperature in °C : 23.5

Humidity in RH % : 50

Standards used :

SI No.	Nomenclature	Make	Model	SI No/ID.No.	Certificate No.	Validity
1	Digital Tachometer	Lutron	DT 2236	L - 510221	SANS/54486/18	14 Mar 2019


Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By



* Tejas
(Calibration Engineer)

Checked By


Annapoorna M P
(Calibration Engineer)



Authorised By


Shreyas B
(Lab Incharge)



Results :-

Department: Laboratory

Sl. No.	DUC Set At	STD Reading in rpm	Measurement Uncertainty \pm in rpm
1	Fixed	29.0	0.58


Conclusion Remarks :-

1. Measurement uncertainty reported is at 95% confidence level.

Calibrated By



x Tejas
(Calibration Engineer)

Checked By


Annapoorna M P
(Calibration Engineer)



Authorised By


Shreyas B V
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At :100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Address : M/s. **Transcaal Engineers India Pvt. Ltd**
No. 96/1-1. Shubhashree (At:First Floor) PO. East Park Road. Ninth CRs
Between Sampige Road & Margosa Road, Malleswaram Bangalore Karnataka 560003

Customer's Reference: **SRF No.: TSC/18-19/4715** Dated :27 Jul 2018

Calibration Certificate	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/4715-2	31 Jul 2018	31 Jul 2019	1 of 2

Details of device under calibration (DUC):

DUC :Dead Weights	Calibration Procedure No. : TSC/CAL/605
Make :---	No. of Pages : 2
Range :20kgs	DUC Received :27 Jul 2018
SI No. :---	DUC Condition on receipt : Satisfactory
ID No. :---	Cal At : Mechanical Lab, Transcal

Environmental Conditions:Temp:(20 ± 2)°C, Relative Hum.:(40 to 60)%, Atmos.Pressure:911.6mbar

Standards used:

Sl. No.	Nomenclature	Make	ID No.	Traceable to / Cert No.	Validity
1	Set of Weights	Bhavani	TSC/MECH-79A to J	Transcal/TSC/18-19/INH/MECH-79	14 Aug 2019

Note:

- This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements
- Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal
- Errors if any, in this Cert. shall be brought to notice within 45 days from the date of this Certificate
- Measurement Uncertainty reported is at approximately 95 % confidence level with k=2 ; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated
- Calibration of the DUC are traceable to National/International Standards
- Corrections/erasing, invalidate the Calibration Cert. - exception to the "Final Page or Part of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control)
- Nabl-133 guidelines are adopted for use of NABL symbol.
- Paper Report contains data on both sides as Environment friendly measures.
- In Result Sheets, "Pass" indicates measured readings are within specification limit, "Fail" Indicates measured readings

Calibrated By

Jyoti C
(Calibration Engineer)

Checked By

Chitra
Chitrangadha P R
(Calibration Engineer)



Authorised By

Shreyas B
(Lab Incharge)



SYSCON CALIBRATION CENTRE PVT. LTD.

Service With Precision

An ISO 17025 Accredited Laboratory



CALIBRATION CERTIFICATE

Certificate No. **SCCPLCRTN201800655**

Date of Receipt	23-07-2018	Calibration Date	23-07-2018	Calibration Due Date	22-07-2019
-----------------	------------	------------------	------------	----------------------	------------

Customer Name & Address:

Transcaal Engineers India Pvt. Limited.
No. 96/1-1, Shubhashree (At: first floor), East Park Road,
9th cross, Between Sampige and Margosa road, Malleswaram
Bangalore - 560003

Customer Reference:

Purchase Order No: 18-19-016
SCCPL Crd/or No.: OC201800241
Instrument SEAL : Intact/Not Intact
Physical Condition: DAMAGED: Yes/No ✓

Details of Equipment:

Nomenclature of Equipment	Make	Model No.	Serial No.	Capacity	Resolution
Gas Flow Analyzer	Fluke	V,T Mobile	3008037	150 L/min	0.01 ml/min

Calibration Related:

Calibration Parameter: Flow Reference Standard: -
Calibration Procedure: SCCPL/SOPFL/002 Mode of Calibration: Air Flow
Calibration Location: SCCPL, Bangalore Local 'g' Value: 9.77997629 m/Sec²

Standards used for calibration Method Used : By Comparison

Description	Traceability & Validity
Fukuda, Laminar Flow element MFU-110-200LM-O, 3N1048,20-200NL/min	CA 452 1805 139 & 30-05-2020
Laminar Flow element,Fukuda, MFU-110-50LM-O, 51150095,Capacity:5-50NL/min	CA 31 1707 160 & 20-07-2019

Environmental Condition:

Temperature: (23 ± 1)°C

Humidity: (50± 10) %RH

Calibration Result Summary: As Per Enclosed Results

- Note: 1. The evaluation of uncertainty of measurement includes the resolution of UUC (unit under calibration).
2. The calibration results relate only to the item calibrated.
3. The results are valid only under the stated conditions of calibration.
4. This calibration certificate shall not be reproduced, except in full without written approval of SCCPL.
5. The calibration of the DUC is traceable to National / International Standards, the measurements are in SI units and process is in line with the Quality Manual and conforming to ISO/IEC 17025: 2005.

CALIBRATED BY		APPROVED BY (Any one required)	
Sign:	Sign:	Sign:	Sign:
Name: Vasantha Kumar	Name: K R Ramu	Name: Shivalaela 23/7/18	Name: R.Vasantha Kumar
Design:	Design: DGM -Quality	Design: Dy. Manager - Quality	Design: Sr.Calibration Officer

Clause 5.10.2, Format SCCPL/QFM/013 R3

27-07-2018

CERTIFICATE OF CALIBRATION					
Issued To	SARADA KRISHNA HOMEOPATHIC MEDICAL COLLEGE & HOSPITAL, KULASEKHARAM				Page 1 of 4
Description	SUCTION MACHINE	Certificate No	TEPL/18-19/1146-24		
Manufacturer	LIFE CARE	Department	CASUALTY		
Model	GOLEY	Cal Procedure	TEPL-CAL-SM-01		
Serial No.	NA	Calibration At	ON SITE		
Asset No.	NA	DUC Condition on Receipt	SATISFACTORY		
Environmental Conditions	Temperature: (30) °C		Relative Humidity: (76)%		
CALIBRATION STANDARDS					
Test Equipment	Manufacturer	Model	Serial No.	Cal Date	Due Date
GAS FLOW ANALYZER	FLUKE	VT MOBILE	3008037	23.07.2018	22.07.2019
ELECTRICAL SAFETY ANALYZER	FLUKE	ESA 615	2839286	19.12.2017	19.12.2018
Calibration Status	Pass <input checked="" type="checkbox"/>		Fail <input type="checkbox"/>		Report Attached
Calibration Date: 08-07-2018			Calibration Due Date: 08-07-2019		

The above mentioned instrument has been calibrated using standards, manufacturer recommended protocols, using equipments having traceability to National & International Standards. Devices for which there are no NIST calibration standards are measured against in-house performance standard using accepted test procedures. Partial reproduction of the certificate is not permitted

Calibration Engineer: JITESH



Approved By: _____

DIRECTOR



SUCTION MACHINE TEST REPORT

Certificate No. TEPL/18-19/1146-24

Page 2 of 4

TEST EQUIPMENT USED : Gas Flow Analyzer


Manufacturer	Fluke
Model	VT MOBILE
Serial Number	3008037
Calibration Date	23-07-2018
Calibration Due	22-07-2019

TOURNIQUET TEST

Sl.No	DUT		TEST GADGET	Tolerance
	Set Value	Indicated Value	Measured Value in mmHg	
1	- 200	- 200	- 198	± 10%
2	- 400	- 400	- 397	± 10%
3	- 600	- 600	- 595	± 10%

Remarks: UNIT TEST PASS

Engineers Sign:



Date: 08-07-2018

Operator ID: TEPL
 Calibration Tech : 22881
 Calibration Date : 19/12/2017
 Firmware Version : 2.12
 Serial Number: 2839286

Template Information

Template Name:	60601 3rd Generic Device	Standard:	IEC60601-1-3rd Ed
Pause after Power ON:	NO	Pause before Power OFF:	NO
Power ON delay:	2	Power OFF delay:	0
Test Speed:	NORMAL	Test Mode:	STEP BY STEP
Halt on Test Failure:	YES	Multi PE Test:	YES
Include Time:	YES	Multi Resstore:	WORST/LAST
Insulation Resistance \geq 500V		Reverse Polarity:	YES
Multi Enclosure Test :	YES		


PLC Configuration-Applied part setup

AP Name: AP Type: AP Num:

ESA615 Test Results

Test Name	Value	High Limits	Low Limits	Status
Protective Earth Resistance	0.14 Ohm	0.2	-	P
Insulation Resistance				P
Mains to Protective Earth	999 MOhm	-	-	2 P
Mains to Non-Earth Accessible Conductive Part	999 MOhm	-	-	7 P
Mains to Applied Parts	999 MOhm	-	-	10 P
Applied Parts to Non-Earth Accessible Conductive	999 MOhm	-	-	70 P
Mains Voltage				P
Live to Neutral	232.1 V	-	-	P
Neutral to Earth	1.3 V	-	-	P
Live to Earth	232.3 V	-	-	P
Equipment Current	0.1 A	-	-	P
Earth Leakage Current				P
Normal Condition	27.3 uA-OPEN	5000	-	P
Enclosure Leakage Current				P
Normal Condition	2.5 uA-OPEN	100	-	P

Open Earth	2.4 uA-OPEN	500 -	P
Open Earth	2.3 uA-OPEN	500 -	P
Open Earth	2.5 uA-OPEN	500 -	P
Earth Leakage Current			P
Open Neutral	36.1 uA-OPEN	10000 -	P
Enclosure Leakage Current			P
Open Neutral	3.2 uA-OPEN	500 -	P
Earth Leakage Current			P
Open Neutral- Reversed Mains	31.1 uA-OPEN	10000 -	P
Enclosure Leakage Current			P
Open Neutral- Reversed Mains	3.4 uA-OPEN	500 -	P
Open Neutral- Reversed Mains	3.2 uA-OPEN	500 -	P
Open Neutral- Reversed Mains	3.1 uA-OPEN	500 -	P
Earth Leakage Current			P
Normal Condition- Reversed Mains	16.1 uA-OPEN	5000 -	P
Enclosure Leakage Current			P
Normal Condition- Reversed Mains	0.9 uA-OPEN	100 -	P
Normal Condition- Reversed Mains	0.8 uA-OPEN	100 -	P
Normal Condition- Reversed Mains	2.4 uA-OPEN	100 -	P
Normal Condition- Reversed Mains	2.2 uA-OPEN	100 -	P
Normal Condition- Reversed Mains	2.0 uA-OPEN	100 -	P
Normal Condition- Reversed Mains	2.1 uA-OPEN	100 -	P
Normal Condition- Reversed Mains	2.1 uA-OPEN	100 -	P
Open Earth- Reversed Mains	2.7 uA-OPEN	500 -	P
Open Earth- Reversed Mains	2.5 uA-OPEN	500 -	P
Open Earth- Reversed Mains	2.2 uA-OPEN	500 -	P
Open Earth- Reversed Mains	2.3 uA-OPEN	500 -	P
Open Earth- Reversed Mains	2.3 uA-OPEN	500 -	P
Open Earth- Reversed Mains	2.2 uA-OPEN	500 -	P
Open Earth- Reversed Mains	2.2 uA-OPEN	500 -	P





TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G.Flr), Door No. At : 100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Transcaal Engineers India Pvt. Ltd
No. 96/1-1, Shubhashree (At:First Floor) PO. East Park Road. Ninth CRs Between
Sampige Road & Margosa Road, Malleswaram Bangalore Karnataka 560003

Customer's Reference : SRF No. : TSC/18-19/4416 Dated : 19 Jul 2018

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/4416-6	23 Jul 2018	23 Jul 2019	1 of 2

Details of device under calibration		Transcal ID : TSC17132	
Nomenclature	: Digital RTD Calibrator	No. of Pages	: 2
Make	: Metravi	Cal Procedure No.	: TSC/CAL/323
Model/Range	: 03	DUC Received	: 19 Jul 2018
SI No.	: 995117115	DUC Condition on Receipt	: Satisfactory
ID No.	: TEPL/16/RTC-021	Cal At	: Electro Technical Lab

Environmental Conditions : Temperature in °C : 25±2 Humidity in RH % : 45-75

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Digital Multimeter	Keysight	34461A	MY57206365	TSC/17-18/INH-ECAL-164	06 Nov 2018
2	Multi-Product Calibrator	Fluke	5520A	9175011	CR/PCAL/49523	31 Oct 2018

Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By

Dhanya M M
(Calibration Engineer)

Checked By

Subramanya B A
(Calibration Engineer)



Authorised By

Ram Rathan G S
(Director Operations)



TransCal

Measurement to Perfection...



Main Bld. Premises: Centenary Building (G. Flr), Door No. At :100, W. Park Rd.,
Between Sampige Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin- 560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Transcaal Engineers India Pvt. Ltd
No. 96/1-1, Shubhashree (At:First Floor) PO. East Park Road, Ninth CRs Between
Sampige Road & Margosa Road, Malleswaram Bangalore Karnataka 560003

Customer's Reference : SRF No. : TSC/18-19/2023 Dated : 04 May 2018

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/2023-1	07 May 2018	07 May 2019	1 of 2

Details of device under calibration		Transcal ID	
Nomenclature	: Steel Weight	No. of Pages	: 2
Make	: --	Cal Procedure No.	: TSC/CAL/605
Model/Range	: 500 g - 10 kg	DUC Received	: 04 May 2018
SI No.	: NA	DUC Condition on Receipt	: Satisfactory
ID No.	: TEPL/17/WGH-028	Cal At	: Mechanical Lab

Environmental Conditions : Temperature in °C : 21.5 Humidity in RH % : 52

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	500g - 1kg E1 Class Weights	BI	--	TSC/Mech-82A to 82H	15060280/D5.01/C-235	12 Jan 2019
2	2kg - 20kg E1 Class Weights	BI	--	TSC/Mech-82A to 82H-1	15060280/D5.01/C-204	10 Nov 2018

Note:

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part' of this Report- provided for incorporation of additional data. To be filled by customer authorized signatory and not under calibration laboratory control.
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By

Jothi C
(Calibration Engineer)

Checked By

Chirangadha P R
(Calibration Engineer)



Authorised By

Shreyas B V
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At :100, W.Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Transcaal Engineers India Pvt. Ltd
No. 96/1-1. Shubhashree (At. First Floor) PO. East Park Road. Ninth CRs Between
Sampige Road & Margosa Road, Malleswaram Bangalore Karnataka 560003

Customer's Reference : SRF No. : TSC/18-19/ 1972 Dated : 03 May 2018

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/ 1972-2	14 May 2018	14 May 2019	1 of 7

Details of device under calibration		Transcal ID : TSC120882	
Nomenclature	: Defibrillator Analyzer	No. of Pages	: 7
Make	: Fluke	Cal Procedure No.	: PROCEEDURE1
Model/Range	: Impulse 7000 DP	DUC Received	: 03 May 2018
SI No.	: 2985063	DUC Condition on Receipt	: Satisfactory
ID No.	: TEPL/15/DBA-004	Cal At	: RF Lab

Environmental Conditions : Temperature in °C : 25±2

Humidity in RH % : 45-75

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Digital Multimeter	Keysight	34461A	MY57206365	TSC/17-18/INH-ECAL-164	06 Nov 2018
2	Multi-Product Calibrator	Fluke	5520A	8130001	TSC/17-18/INH-ECAL-01	03 Oct 2018
3	Universal Frequency Counter	Phillips	PM 6669	BQ 1953	TSC/17-18/INH/ECAL-05	23 Dec 2018
4	Digital Oscilloscope	Tektronix	M504104	C000335	TSC/17-18/INH/RF/HV-07	19 Feb 2019

Calibrated By

Geetha K S
(Calibration Engineer)

Checked By

Malikarjun V R
(Sr. Calibration Engineer)



Authorised By

IS Prasad
(General Manager)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampige Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin - 560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. **Transcaal Engineers India Pvt. Ltd**
No. 96/1-1. Shubhashree (At:First Floor) PO. East Park Road. Ninth CRs Between
Sampige Road & Margosa Road, Malleswaram Bangalore Karnataka 560003

Customer's Reference : **SRF No. : TSC/18-19/ 1972** Dated : 03 May 2018
Supplement to original certificate: TSC/18-19/1972-1 Dated 10 Aug 2018

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/18-19/ 1972-1A	18 May 2018	18 May 2019	1 of 7

Details of device under calibration Transcal ID : TSC45089

Nomenclature : Vital Signs Simulator	No. of Pages : 7
Make : Fluke	Cal Procedure No. : TSC/CAL/323
Model/Range : Prosim 4	DUC Received : 03 May 2018
SI No. : 2989669	DUC Condition on Receipt : Satisfactory
ID No. : TEPL/15/VSS-001	Cal At : RF Lab

Environmental Conditions : Temperature in °C : 25±2

Humidity in RH % : 45-75

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Digital Multimeter	Keysight	34461A	MY57206365	TSC/17-18/INH-ECAL-164	06 Nov 2018
2	Multi-Product Calibrator	Fluke	5520A	8130001	TSC/17-18/INH-ECAL-01	03 Oct 2018
3	Universal Frequency Counter	Phillips	PM 6669	BQ 1953	TSC/17-18/INH/ECAL-05	23 Dec 2018
4	Digital Oscilloscope	Tektronix	M504104	C000335	TSC/17-18/INH/RF/HV-07	19 Feb 2019

Calibrated By

Geetha K S
Geetha K S
(Calibration Engineer)

Checked By

Mallikarjun V R
Mallikarjun V R
(Sr. Calibration Engineer)



Authorised By

IS Prasad
IS Prasad
(General Manager)



Tektronix India Pvt Ltd
MBC Centre, Ground Floor,
Plot no 143/A, Ghodbunder Road
Opp. Cinewonder Cinema
Kapurbawadi, Thane-400607
INDIA

Phone No:91-22-32285210

CERTIFICATE OF TRACEABLE CALIBRATION

Certificate No: 1851809-4-SPOT LIGHT-2989669-1

Contract/PO No:

Customer:

Transcaal Engineers India Pvt.Ltd.
Bldg no.96/1-1,Shubhashree,1st Floor,
Near Post Office, East Park Road,9th Cross,
Malleswaram Road, Bangalore-560003

Model:	SPOT LIGHT
Serial No.:	2989669
Manufacturer:	Fluke Biomedical
Description:	Sp02 FUNCTIONAL TESTER
Site of Calibration:	Service Center
Calibration Interval Source:	Tektronix recommended
Cal Date:	19-Feb-2018
Due Date:	19-Feb-2019
Calibration Interval:	12Months
Temperature(25±2°C):	23.0°C
Humidity (40-70%RH):	45 %
Received Date:	19-Feb-2018

Tektronix certifies that the performance of the above instrument has been verified using test equipment of known accuracy which are traceable to National & International Standards. This certificate shall not be reproduced except in full , with the written approval of the calibration facility.



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At :100, W. Park Rd.,
Between Sampighe Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. **Transcaal Engineers India Pvt. Ltd**
No. 96/1-1. Shubhashree (At:First Floor) PO. East Park Road. Ninth CRs Between
Sampige Road & Margosa Road, Malleswaram Bangalore Karnataka 560003

Customer's Reference : SRF No. : TSC/17-18/ 10262 Dated : 18 Feb 2018

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/17-18/ 10262-1	18 Feb 2018	18 Feb 2019	1 of 2

Details of device under calibration		Transcal ID	: TSC112545
Nomenclature	: Pressure Calibrator	No. of Pages	: 2
Make	: R&D Instruments	Cal Procedure No.	: TSC/CAL/601&602
Model/Range	: EPC40	DUC Received	: 18 Feb 2018
SI No.	: 19591702	DUC Condition on Receipt	: Satisfactory
ID No.	: TEPL/16/PCI-017	Cal At	: Mechanical Lab

Environmental Conditions : Temperature in °C : 23.1

Humidity in RH % : 54

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Reff Pressure Gauge	Fluke	2700G	2448199	2017-18/CFC/2582/1	12 Feb 2019
2	Dead Weight Tester	Fluke	P3125XT-BAR	70655	2017-18/CFC/2582/2	17 Feb 2020

Note :

1. This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
2. Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
3. Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
4. Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
5. Calibration of the DUC are traceable to National/International Standards
6. Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part. of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
7. In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
8. Consider Model or Range whichever is applicable.
9. Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By

Kouathir

X Jothi C
(Calibration Engineer)

Checked By

Manjanath D
Manjanath D
(Calibration Engineer)

Authorised By



Shreyas B V
Shreyas B V
(Lab Incharge)



TransCal

Measurement to Perfection...



Main Bld. Premises : Centenary Building (G. Flr), Door No. At : 100, W. Park Rd.,
Between Sampige Road And Margosa Rd., 10th Crs., Malleswaram, Bangalore City, Pin-560003

CC-2231

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2005

CALIBRATION CERTIFICATE

Customer Name & Add. : M/s. Transcaal Engineers India Pvt. Ltd
No. 96/1-1. Shubhashree (At:First Floor) PO. East Park Road. Ninth CRs Between
Sampige Road & Margosa Road, Malleswaram Bangalore Karnataka 560003

Customer's Reference : SRF No. : TSC/17-18/8950 **Dated :** 19 Dec 2017

Calibration Certificate Number	Calibrated On	Recommended Calibration Due	Page Number
TSC/17-18/8950-1	19 Dec 2017	19 Dec 2018	1 of 3

Details of device under calibration		Transcal ID	
Nomenclature : Electrical Safety Analyzer		Transcal ID : TSC43830	
Make : Fluke		No. of Pages : 3	
Model/Range : ESA 615		Cal Procedure No. : TSC/CAL/312	
SI No. : 2839286		DUC Received : 19 Dec 2017	
ID No. : --		DUC Condition on Receipt : Satisfactory	
		Cal At : Electro Technical Lab	

Environmental Conditions : Temperature in °C : 25±2 Humidity in RH % : 45-75

Standards used :

SI No.	Nomenclature	Make	Model	SI No./ID.No.	Certificate No.	Validity
1	Decade Megohm Box	Vaiseshika	8400 HV	6374	TSC/17-18/INH/ECAL/23	04 Dec 2018
2	Decade Resistance Box	Vaiseshika	7400	5840	TSC/17-18/INH-ECAL-36	04 Dec 2018
3	Dig. Multimeter	Agilent	34401A	US36038748	TSC/16-17/INH/ECAL-60	24 Jan 2018
4	Digital Oscilloscope	Tektronix	MSO4104	C000335	TSC/16-17/INH/RF/HV-07	20 Feb 2018

Note :

- This Calibration Certificate relates only to the above DUC & Reported results are valid at the time of and under the stated conditions of measurements.
- Partial Publication/ reproduction of this Certificate in any form is not permitted without the written consent of Transcal.
- Errors if any, in this Certificate shall be brought to notice within 45 days from the date of this Certificate
- Measurement Uncertainty reported is at approximately 95 % confidence level with k=2; Units of Measurement results & Measurement Uncertainty are same as that of range selected - Unless otherwise indicated.
- Calibration of the DUC are traceable to National/International Standards
- Corrections/erasing, invalidate the Calibration Certificate- exception to the 'Final Page or Part of this Report- provided for incorporation of additional data(To be filled by customer authorized signatory and not under calibration laboratory control).
- In Result Sheets, 'Pass' indicates measured readings are within specification limit, 'Fail' Indicates measured readings are out of specification limit & '-' indicates no specification limit furnished.
- Consider Model or Range whichever is applicable.
- Nabl-133 guidelines are adopted for use of NABL symbol.

Calibrated By

Satheesh Kumar C R
(Calibration Engineer)

Checked By

Vijay Kumar N N
(Sr. Calibration Engineer)



Authorised By

IS Prasad
(General Manager)